



IN-HOUSE CALIBRATION CERTIFICATE

Name of the Equipment :- Centrifuge Machine	
Equipment I.D :- UHM/ICTC/centrifuge/01	
Location :- HTL	Working Range :- 0-5000
Manufacturer :- REMI	
Calibrated on :- 23/04/2022	Next Calibration due :- 22/04/2022

Method :- Traceability Certificate

GSVMC/Micro/HTL/Digital Tachometer/01 SRF No. & Date: 211119A/N Dt. 03/12/2021
Standard: Line Seiki/TM5010EK-039, Godrej, M-210914-16-1 Dt. 14/09/2022

Test Condition: Working


Room Temperature: 24°C


RH: 34%

Instrument	Reading					Mean
	1	2	3	4	5	
Tachometer (GSVMC/Micro/HTL/Tachometer/01)	3008	3010	3005	3009	3015	3009
Centrifuge 02 (GSVMC/Micro/HTL/centrifuge/02)	3000	3000	3000	3000	3000	3000

Acceptable Variation: +/- 10% with standard Tachometer

Remarks: Instrument has passed the acceptance criteria (Yes/No).....Yes


Checked by
Lab Technician
HCTS/HTL
Dept. of Microbiology
G.S.V.M. Medical College
Kanpur


Verified by
Technical Manager
HIV Testing Laboratory (HTL)
Dept. of Microbiology
G.S.V.M. Medical College
Kanpur