

**CORRECTIVE ACTION STATUS:**

Problem	<b><u>Sodium ISE Direct result exceeding the SDI range in EQAS Jan 2022 results.</u></b> <b><u>SDI Reading : 4.48</u></b> <b><u>Result Date: 17 -01-2022</u></b>
Corrective Action Plan	<b><u>Equipment to be checked and repaired by the authorized service team within 30 days.</u></b> <b><u>Instrument model: Innolyte</u></b>
Estimated Correction Time	<b><u>30 days</u></b>
Status of Correction	<b><u>Completed</u></b>
Evidence of Correction	<b><u>Labx Instrument Repair Report dated 20-01-2022</u></b>
Desired Result Achieved Status	<b><u>Yes</u></b>
Evidence of sDesired Result	<b><u>SDI range in EQAS Feb 2022 results.</u></b> <b><u>SDI Reading : 1.43</u></b> <b><u>Result Date: 12-02-2022</u></b>