

## CALIBRATION CERTIFICATE

SANS/FT/01(7.8/01)	Page No : 1 of 2
Certificate/ULR No : CC287921000004492F	Date of issue : 02.09.2021
Date of Calibration : 01.09.2021	Due date : 31.08.2022

### Customer Details:

M/s.VAIDYANATHESHWARA INSTRUMENTS No.301/A, 9 th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore-560 096.	SRF No : 13901 Date of Receipt : 31.08.2021 Cond of Receipt : Satisfactory Calibrated at : LAB
--	---


### Details of UUC

Instrument Name : Digital Tachometer (Contact) & ( Non - Contact)	ID No. : VI/ME/DT/01
Make : SENTRY	
Model : ST723	
S.No : SN1323004225	
Cal . Range : 100 to 8000 rpm (Contact) & 12 to 90000 rpm (Non- Contact)	
Resolution : 0.1/1 rpm	
Accuracy : ---	


<b>Environmental Details</b> Ambient Temperature : 24.1°C Relative Humidity : 50.3%RH	<b>Method of Calibration</b> Comparison Method	<b>Work Instruction</b> : SANS/M/L-02 <b>Discipline</b> : Mechanical <b>Group</b> : Acceleration and Speed
---	---	--

### Details of Standards Used

Nomenclature	Sl.No	Certificate No	Validity	Traceability
Digital Tachometer	168-0093	FCRI/EQL/20-21/327	07.10.2022	National standard
Digital Tachometer	109-0591V	HT/CC/200915-27/001	14.09.2021	National Standard
RPM Source	RPMC11160798	---	---	---

  
**K.Sudhakar**  
Calibration Engineer



  
**Anand B**  
Authorised Signatory

**Remarks :** 1. The recalibration intervals are determined based on user requirements. 2. Meters/Equipment used for calibration are traceable to National & International Standards. 3. This Certificate refers only to particular item(s) submitted for calibration. 4. The Calibration results reported are valid at the time of and under the stated conditions of the measurements. 5. The calibration certificate shall not be reproduced without prior permission from the laboratory. 6. All precautions have been taken to avoid any error or omission while calibrating the liability instruments and issuing this certificate. However the laboratory shall not be liable for any loss or that may arise to any part in this regard. 7. UUC: Unit Under Calibration, C.L: Confident Level.