

**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021


**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

**Qualitative Immunoassay,Dengue NS 1, -**

	<b>N</b>
<span style="display: inline-block; width: 15px; height: 15px; background-color: #9933cc; border: 1px solid black; margin-right: 5px;"></span> All Methods	60
<span style="display: inline-block; width: 15px; height: 15px; background-color: #008080; border: 1px solid black; margin-right: 5px;"></span> Immunochromatography	51
<span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black; margin-right: 5px;"></span> Manual	62

<b>Your Result</b>	Negative
<b>Result Expected</b>	Neg

**Note :** # is an indication that this parameter has not been evaluated



**Reported Results Legends :-**

Pos	: Positive
Eq	: Equivocal
BL	: Borderline
Neg	: Negative

**Comments :**



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 2 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance. Review for systematic error



**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 2  
 Sample Date : 03-03-2021


**Qualitative Immunoassay,HBsAg, Nil**

	<b>N</b>
<span style="color: purple;">■</span> All Methods	92
<span style="color: teal;">■</span> Immunochromatography	54
<span style="color: maroon;">■</span> Manual	43

<b>Your Result</b>	Non-reactive
<b>Result Expected</b>	NR

**Note** : # is an indication that this parameter has not been evaluated

**Frequency Histogram**



**Reported Results Legends :-**

R : Reactive  
 NR : Non-reactive  
 Eq : Equivocal

**Comments :**



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
- Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
- b) Reported on 3 samples:
- All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 2  
 Sample Date : 03-03-2021


**Qualitative Immunoassay,HCV, -**

<span style="display: inline-block; width: 15px; height: 15px; background-color: #9933cc; border: 1px solid black;"></span>	All Methods	N
<span style="display: inline-block; width: 15px; height: 15px; background-color: #008080; border: 1px solid black;"></span>	Immunochromatography	80
<span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black;"></span>	Manual	46
		38

<b>Your Result</b>	Non-reactive
<b>Result Expected</b>	NR

**Note :** # is an indication that this parameter has not been evaluated

Frequency Histogram



<b>Reported Results Legends :-</b>	
R	: Reactive
NR	: Non-reactive
Eq	: Equivocal

**Comments :**



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 2  
 Sample Date : 03-03-2021

**Qualitative Immunoassay,HIV Ab, -**


	<b>N</b>
<span style="display: inline-block; width: 15px; height: 15px; background-color: #9933cc; border: 1px solid black;"></span> All Methods	39
<span style="display: inline-block; width: 15px; height: 15px; background-color: #008080; border: 1px solid black;"></span> Immunochromatography	37
<span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black;"></span> Manual	33

**Your Result** Reactive

**Result Expected** #

**Note** : # is an indication that this parameter has not been evaluated

Frequency Histogram



**Reported Results Legends :-**

- R : Reactive
- NR : Non-reactive
- Eq : Equivocal
- Indet : Indeterminate
- HIV-1 : HIV-1 Ab Reactive
- HIV-2 : HIV-2 Ab Reactive



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

**Comments :**

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021


Cycle No : 210103  
 Sample No : 2  
 Sample Date : 03-03-2021

**Qualitative Immunoassay,RPR (Syphilis Serology), -**

	<b>N</b>
<span style="color: purple;">■</span> All Methods	72
<span style="color: teal;">■</span> Immunochromatography	28
<span style="color: darkred;">■</span> Manual	74

<b>Your Result</b>	Non-reactive
<b>Result Expected</b>	NR

**Note** : # is an indication that this parameter has not been evaluated



**Reported Results Legends :-**

R	: Reactive
NR	: Non-reactive
WR	: Weak Reactive

**Comments :**



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 2  
 Sample Date : 03-03-2021


**Qualitative Immunoassay, Typhidot IgM, -**

<p><span style="color: purple;">■</span> All Methods</p> <p><span style="color: teal;">■</span> Immunochromatography</p> <p><span style="color: darkred;">■</span> Manual</p>	<p>N</p> <p>58</p> <p>59</p> <p>60</p>
---	--

<b>Your Result</b>	Negative
<b>Result Expected</b>	Neg

**Note** : # is an indication that this parameter has not been evaluated

Frequency Histogram



<b>Reported Results Legends :-</b>	
Pos	: Positive
Neg	: Negative

**Comments :**



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 2  
 Sample Date : 03-03-2021


**Qualitative Immunoassay,Widal test - ParaTyphi A-H antigen (AH), -**

	<b>N</b>
<span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black;"></span> All Methods	71
<span style="display: inline-block; width: 15px; height: 15px; background-color: #008080; border: 1px solid black;"></span> Slide Agglutination	49
<span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black;"></span> Manual	73

<b>Your Result</b>	No Agglutination
<b>Result Expected</b>	NoA

**Note** : # is an indication that this parameter has not been evaluated

Frequency Histogram



**Reported Results Legends :-**

NoA	: No Agglutination
1:80	: 1:80
1:160	: 1:160
1:320	: 1:320
1:640	: 1:640
>1:640	: >1:640



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

**Comments :**

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 2  
 Sample Date : 03-03-2021


**Qualitative Immunoassay,Widal test - ParaTyphi B-H antigen (BH), -**

<p><span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black; margin-right: 5px;"></span> All Methods</p> <p><span style="display: inline-block; width: 15px; height: 15px; background-color: #008080; border: 1px solid black; margin-right: 5px;"></span> Slide Agglutination</p> <p><span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black; margin-right: 5px;"></span> Manual</p>	<p>N</p> <p>71</p> <p>49</p> <p>73</p>
--	--

<b>Your Result</b>	No Agglutination
<b>Result Expected</b>	NoA

**Note** : # is an indication that this parameter has not been evaluated

Frequency Histogram



**Reported Results Legends :-**

NoA	: No Agglutination
1:80	: 1:80
1:160	: 1:160
1:320	: 1:320
1:640	: 1:640
>1:640	: >1:640



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

**Comments :**

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error





**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
Report Date : 19/03/2021

Cycle No : 210103  
Sample No : 2  
Sample Date : 03-03-2021


**Qualitative Immunoassay,Widal test - Typhi H antigen (TH), -**

<p><span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black; margin-right: 5px;"></span> All Methods</p> <p><span style="display: inline-block; width: 15px; height: 15px; background-color: #008080; border: 1px solid black; margin-right: 5px;"></span> Slide Agglutination</p> <p><span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black; margin-right: 5px;"></span> Manual</p>	<p>N</p> <p>81</p> <p>56</p> <p>83</p>
--	--

<b>Your Result</b>	No Agglutination
<b>Result Expected</b>	NoA

**Note** : # is an indication that this parameter has not been evaluated

Frequency Histogram



**Reported Results Legends :-**

NoA	: No Agglutination
1:80	: 1:80
1:160	: 1:160
1:320	: 1:320
1:640	: 1:640
>1:640	: >1:640



**Analyte Methodwise Qualitative Result Report**

**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

**Comments :**

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Analyte Methodwise Qualitative Result Report**

Lab No.: 455  
 Lab Name : MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
 Report Date : 19/03/2021

Cycle No : 210103  
 Sample No : 2  
 Sample Date : 03-03-2021


**Qualitative Immunoassay,Widal test - Typhi O antigen (TO), -**

	<b>N</b>
<span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black; margin-right: 5px;"></span> All Methods	81
<span style="display: inline-block; width: 15px; height: 15px; background-color: #008080; border: 1px solid black; margin-right: 5px;"></span> Slide Agglutination	56
<span style="display: inline-block; width: 15px; height: 15px; background-color: #800000; border: 1px solid black; margin-right: 5px;"></span> Manual	83

<b>Your Result</b>	No Agglutination
<b>Result Expected</b>	NoA

**Note** : # is an indication that this parameter has not been evaluated

Frequency Histogram



**Reported Results Legends :-**

NoA	: No Agglutination
1:80	: 1:80
1:160	: 1:160
1:320	: 1:320
1:640	: 1:640
>1:640	: >1:640



**Analyte Methodwise Qualitative Result Report**


**Lab No.:** 455  
**Lab Name :** MEDICENTRE SONOGRAPHY & CLINICAL LAB, DUNGERPUR  
**Report Date :** 19/03/2021

**Cycle No :** 210103  
**Sample No :** 2  
**Sample Date :** 03-03-2021

**Comments :**

Interpretation when an analyte is reported for multiple samples in a Round:

- a) Reported on 2 samples:
  - Two out of two pass - Acceptable performance
  - One out of two pass - Review for random error & take action if appropriate
  - Both failed - Unacceptable performance, Review for systematic error
  
- b) Reported on 3 samples:
  - All pass - Acceptable performance
  - Two out of three pass - Review for random error & take action if appropriate
  - One out of three pass - Unacceptable performance, Review for systematic error
  - All failed - Unacceptable performance, Review for systematic error



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070



PC-1022