



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

Certificate No:	22900002208	Issued On:	28/04/2022
ULR No:	CC214622900002208F	Calibration Date:	27/04/2022
Job Identification No:	ICS/C/AS/04/2208	Next Calibration Date:	26/04/2023
Ref. No:	SRF- Dated: 25/04/2022		

<b>CALIBRATED FOR:</b>	M/s Q-4 Health & Wellness Pvt. Ltd. First Floor, A-39, Okhla Phase-I, Delhi-110020.
------------------------	---

#### EQUIPMENT DESCRIPTION

Name	Digital Laboratory Centrifuge, Fitted with Digital RPM Indicator		
Make/Model	Remi / R-8C	Visual Inspection	Ok
Serial No.	ZGIN-19297	I.D. No.	CEN-01
Range	Up to 4400 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

#### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS: Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Measurement Method)

RESULTS: Mechanical Calibration  
MEASUREMENT: (in RPM)


Centrifuge' Reading	Corrected Value Of Tachometer Reading (Average)	Uncertainty At Approx 95% C.L And Coverage Factor K=2
1720	1722.3	± 6
2010	2013.1	± 6
2480	2484.6	± 7
3130	3136.3	± 7
4260	4267.8	± 8

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT \*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
3. Total liability of our Organisation is limited to the invoiced amount.  
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by  Authorised Signatory  
RAJESH KAPOOR  
Operation-Head



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF MICROPIPETTE

<b>Certificate No:</b>	22700006481	<b>Issued On:</b>	28/04/2022
<b>ULR No:</b>	CC214622700006481F	<b>Calibration Date:</b>	27/04/2022
<b>Job Identification No:</b>	ICS/C/MVD/04/6481	<b>Next Calibration Date:</b>	26/04/2023
<b>Ref. No:</b>	SRF- Dated: 25/04/2022		

**CALIBRATED FOR:** M/s Q-4 Health & Wellness Pvt. Ltd.  
First Floor, A-39, Okhla Phase-I,  
Delhi-110020.

### EQUIPMENT DESCRIPTION

<b>Name</b>	Variable Micropipette		
<b>Make/Model</b>	S4 Pipette	<b>Visual Inspection</b>	Ok
<b>Serial No.</b>	---	<b>I.D. No.</b>	MP-01
<b>Range</b>	100 µl - 1000 µl	<b>Least Count</b>	10 µl
<b>Location</b>	---	<b>Calibration Site</b>	In-Lab

### ENVIRONMENTAL CONDITIONS

<b>Temperature</b>	25.0 °C ± 3.0° C	<b>Humidity</b>	50 ± 10 % RH
--------------------	------------------	-----------------	--------------

### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

<b>USED EQUIPMENT DETAILS</b>	Calibrated Balance, Distilled Water & Glass wares
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655-6

**RESULTS: Mechanical Calibration**  
Below Volume is determined at 27°C


U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.1000 (0100 µl)	0.09973 (0099.73 µl)	± 6 µl
0.5000 (0500 µl)	0.49954 (0499.54 µl)	± 6 µl
1.0000 (1000 µl)	0.99929 (0999.29 µl)	± 6 µl

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by   
Authorised Signatory  
RAJESH KAPOOR  
Operation-Head



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF MICROPIPETTE

<b>Certificate No:</b>	22700006482	<b>Issued On:</b>	28/04/2022
<b>ULR No:</b>	CC214622700006482F	<b>Calibration Date:</b>	27/04/2022
<b>Job Identification No:</b>	ICS/C/MVD/04/6482	<b>Next Calibration Date:</b>	26/04/2023
<b>Ref. No:</b>	SRF- Dated: 25/04/2022		

<b>CALIBRATED FOR:</b>	<b>M/s Q-4 Health &amp; Wellness Pvt. Ltd.</b> First Floor, A-39, Okhla Phase-I, Delhi-110020.
------------------------	--

#### EQUIPMENT DESCRIPTION

<b>Name</b>	Variable Micropipette		
<b>Make/Model</b>	Micpette	<b>Visual Inspection</b>	Ok
<b>Serial No.</b>	---	<b>I.D. No.</b>	MP-02
<b>Range</b>	5 µl - 50 µl	<b>Least Count</b>	1 µl
<b>Location</b>	---	<b>Calibration Site</b>	In-Lab

#### ENVIRONMENTAL CONDITIONS

<b>Temperature</b>	25.0 °C ± 3.0° C	<b>Humidity</b>	50 ± 10 % RH
--------------------	------------------	-----------------	--------------

#### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

<b>USED EQUIPMENT DETAILS</b>	Calibrated Balance, Distilled Water & Glass wares
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655-6

**RESULTS: Mechanical Calibration**  
Below Volume is determined at 27°C


U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.005 (05 µl)	0.005061 (0005.061 µl)	± 0.6 µl
0.025 (25 µl)	0.025125 (0025.125 µl)	± 0.6 µl
0.050 (50 µl)	0.050243 (0050.243 µl)	± 0.6 µl

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
**Indian Calibration Services**

Checked by  Authorised Signatory

Operation-Head



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF MICROPIPETTE

<b>Certificate No:</b>	22700006483	<b>Issued On:</b>	28/04/2022
<b>ULR No:</b>	CC214622700006483F	<b>Calibration Date:</b>	27/04/2022
<b>Job Identification No:</b>	ICS/C/MVD/04/6483	<b>Next Calibration Date:</b>	26/04/2023
<b>Ref. No:</b>	SRF- Dated: 25/04/2022		

<b>CALIBRATED FOR:</b>	<b>M/s Q-4 Health &amp; Wellness Pvt. Ltd.</b> First Floor, A-39, Okhla Phase-I, Delhi-110020.
------------------------	--

### EQUIPMENT DESCRIPTION

<b>Name</b>	Variable Micropipette		
<b>Make/Model</b>	S4 Pipette	<b>Visual Inspection</b>	Ok
<b>Serial No.</b>	---	<b>I.D. No.</b>	MP-03
<b>Range</b>	20 µl - 200 µl	<b>Least Count</b>	1 µl
<b>Location</b>	---	<b>Calibration Site</b>	In-Lab

### ENVIRONMENTAL CONDITIONS

<b>Temperature</b>	25.0 °C ± 3.0° C	<b>Humidity</b>	50 ± 10 % RH
--------------------	------------------	-----------------	--------------

### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name	Make	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
Stainless Steel Wire Weights	Weightronics	WMCL/E/2020-05/3364	WMCL	31/05/2020	30/05/2023
Digital Temp. Indicator with Sensor	"Lutron"	22100000001	Indian Calibration Services, Delhi	01/01/2022	01/01/2023

<b>USED EQUIPMENT DETAILS</b>	Calibrated Balance, Distilled Water & Glass wares
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS/CAL/SOP-M03 (Gravimetric Method), ISO 8655-6

**RESULTS: Mechanical Calibration**  
Below Volume is determined at 27°C

U.U.C. Reading (in ml)	Standard Measured Volume (in ml) (Average)	Uncertainty at approx 95% C.L and coverage factor k=2
0.020 (020 µl)	0.02015 (020.15 µl)	± 0.6 µl
0.100 (100 µl)	0.10027 (100.27 µl)	± 0.6 µl
0.200 (200 µl)	0.20043 (200.43 µl)	± 0.6 µl

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\*END OF REPORT\*\*\*

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
**Indian Calibration Services**

Checked by

Authorised Signatory

RAJESH KAPOOR  
Operation-Head



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechcs.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL TEMPERATURE INDICATOR WITH SENSOR

Calibration Certificate No.	22100007061	Issued On	28/04/2022
ULR No.	CC214622100007061F	Calibration Date	27/04/2022
Job Identification No.	ICS/C/T/04/7061	Next Due Date	26/04/2023
Ref. No.	SRF- Dated: 25/04/2022		

#### CALIBRATED FOR:

M/s Q-4 Health & Wellness Pvt. Ltd.  
First Floor, A-39, Okhla Phase-I,  
Delhi-110020.

#### EQUIPMENT DESCRIPTION

Name	Digital Temperature Indicator with Sensor		
Make/Model	Multi	Visual Inspection	Ok
Serial No.	---	I.D. No.	DTI-01
Range	-50°C to 300°C	Least Count	0.1° C
Location	---	Calibration Site	In-Lab

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

#### STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Yudian	ETL/2113/22/C/F/061	FCRI, Palakkad	02-03-2022 to 04-03-2022	04-03-2023

#### USED EQUIPMENT DETAILS

Methanol Bath, Water Bath & Dry Block Calibrator

#### PRINCIPLE/METHODOLOGY OF CALIBRATION:

As per Calibration Procedure No.: ICS-/CAL/SOP-TH02, ITS-90  
(Comparison Method)

#### RESULTS: Thermal Calibration (On Working Range)

U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
-20.0	-20.138	± 0.18
50.0	49.825	± 0.28
100.0	99.748	± 0.41
200.0	199.637	± 0.41
300.0	299.518	± 1.50

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\* END OF REPORT \*\*\*

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
3. Total liability of our Organisation is limited to the invoiced amount.  
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
Indian Calibration Services

Checked by

Authorised Signatory

Operation-Head



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments  
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092  
Phone / Fax : (011) 22453259, 22434362, 9599482307  
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL TEMPERATURE INDICATOR WITH SENSOR

Calibration Certificate No.	22100007062	Issued On	28/04/2022
ULR No.	CC214622100007062F	Calibration Date	27/04/2022
Job Identification No.	ICS/C/T/04/7062	Next Due Date	26/04/2023
Ref. No.	SRF- Dated: 25/04/2022		

<b>CALIBRATED FOR:</b>	M/s Q-4 Health & Wellness Pvt. Ltd. First Floor, A-39, Okhla Phase-I, Delhi-110020.
------------------------	---

### EQUIPMENT DESCRIPTION

Name	Digital Temperature Indicator with Sensor		
Make/Model	Multi	Visual Inspection	Ok
Serial No.	---	I.D. No.	DTI-02
Range	-50°C to 300°C	Least Count	0.1° C
Location	---	Calibration Site	In-Lab

### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Yudian	ETL/2113/22/C/F/061	FCRI, Palakkad	02-03-2022 to 04-03-2022	04-03-2023

<b>USED EQUIPMENT DETAILS</b>	Methanol Bath, Water Bath & Dry Block Calibrator
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS-CAL/SOP-TH02, ITS-90 (Comparison Method)

**RESULTS: Thermal Calibration**  
(On Working Range)



U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
-20.0	-20.151	± 0.18
50.0	49.803	± 0.28
100.0	99.717	± 0.41
200.0	199.592	± 0.41
300.0	299.466	± 1.50

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.

\*\*\* END OF REPORT \*\*\*

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
**Indian Calibration Services**

Checked by:   
Authorised Signatory: 

Operation-Head



# Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechcs.com

## CALIBRATION CERTIFICATE

Page 1 of 1

### CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100007063	Issued On	28/04/2022
ULR No.	CC214622100007063F	Calibration Date	27/04/2022
Job Identification No.	ICS/C/T/04/7063	Next Due Date	26/04/2023
Ref. No.	SRF- Dated: 25/04/2022		

<b>CALIBRATED FOR:</b>	<b>M/s Q-4 Health &amp; Wellness Pvt. Ltd.</b> First Floor, A-39, Okhla Phase-I, Delhi-110020.
------------------------	--

#### EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	DTH-01
Range	Up to 60°C / Up to 99% (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

#### ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

#### STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Yudian	ETL/2113/22/C/F/061	FCRI, Palakkad	02-03-2022 to 04-03-2022	04-03-2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/C/F/506	FCRI, Kerala	30/12/2021 - 04-01- 2022	04/01/2023

<b>USED EQUIPMENT DETAILS</b>	Environmental Chamber
<b>PRINCIPLE/METHODOLOGY OF CALIBRATION:</b>	As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90) (Comparison Method)

#### RESULTS: Thermal Calibration (On Working Range)

U.U.C. Reading		Corrected Value of Standard (Avg.)		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
15.0°C	20 % RH	15.092°C	20.81 %RH	± 0.33°C	± 1.6 % RH
20.0°C	40 % RH	20.129°C	41.23 %RH	± 0.33°C	± 1.6 % RH
30.0°C	75 % RH	30.161°C	76.61 %RH	± 0.33°C	± 1.6 % RH
40.0°C	90 % RH	40.195°C	91.95 %RH	± 0.33°C	± 1.6 % RH

**REMARKS:** The recommended date for next calibration is mentioned, as asked by the customer.

**\*\*\*END OF REPORT\*\*\***

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
  - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
  - Total liability of our Organisation is limited to the invoiced amount.
  - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
  - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
  - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of  
**Indian Calibration Services**

Checked by

Authorised Signatory

Operation-Head