



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL LABORATORY CENTRIFUGE

Certificate No:	22900002208	Issued On:	28/04/2022
ULR No:	CC214622900002208F	Calibration Date:	27/04/2022
Job Identification No:	ICS/C/AS/04/2208	Next Calibration Date:	26/04/2023
Ref. No:	SRF- Dated: 25/04/2022		

CALIBRATED FOR:	M/s Q-4 Health & Wellness Pvt. Ltd. First Floor, A-39, Okhla Phase-I, Delhi-110020.
------------------------	---

EQUIPMENT DESCRIPTION

Name	Digital Laboratory Centrifuge, Fitted with Digital RPM Indicator		
Make/Model	Remi / R-8C	Visual Inspection	Ok
Serial No.	ZGIN-19297	I.D. No.	CEN-01
Range	Up to 4400 RPM	Least Count	10 RPM
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS: Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Measurement Method)

RESULTS: Mechanical Calibration
MEASUREMENT: (in RPM)



Centrifuge' Reading	Corrected Value Of Tachometer Reading (Average)	Uncertainty At Approx 95% C.L And Coverage Factor K=2
1720	1722.3	± 6
2010	2013.1	± 6
2480	2484.6	± 7
3130	3136.3	± 7
4260	4267.8	± 8

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

***END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by  Authorised Signatory 

Operation-Head