



# RAPID DIAGNOSTIC GROUP OF COMPANIES

\* Rapid Diagnostic Pvt. Ltd.      \* Aspen Diagnostics Pvt. Ltd.  
 \* Aspen Laboratories Pvt. Ltd.    \* Deep Meditech Pvt. Ltd.

B-82, G.T.K. Road Industrial Area Delhi - 110 033 (INDIA)  
 Ph.: 27113030, 45201100, Fax: 27413437, 45201198

WEBSITE : www.rapiddiagnostic.in



## TECHNICAL SERVICE / INSTALLATION REPORT

Customer's Name and Address: Hind labs, Gurugram  
Shri Mata Sheela Devi Swarna Jayanti  
Diagnostic center.

Date: 27/07/2021      TSR No.: **2550**

Inst. Model: URI 600

Cont. Person: Dr. VIJAY K. SHARMA

Inst. S. No.: N1900500410348SE

Contact No.: 8800356516      Email: \_\_\_\_\_

Service Status: Warranty  R/R  A.M.C.  C.M.C.  Chargeable  Others

Call Status: **INSTALLATION**  Breakdown  PM  Routine Service  Other

Problem Reported/ Installation Details: \_\_\_\_\_

\* Calibration Problem,  
Calibration Not okay

- Instrument attended and is working satisfactorily  
 Job to be continued  
 Following parts are required to be replaced  
 Following parts have been replaced

Observations: \* Dirty lens

\* Calibration test okay  
after maintenance

Sr. No.	Parts	Old Hologram	New Hologram	Price (Rs.)
1.				
2.				
3.				
4.				
5.				
Total (Rs.)				

Action Taken: \* Cleaned the dirty lens

\* Ran Calibration Strip  
thrice

Customer's Remarks (If any): \_\_\_\_\_

Engineer's Remarks (If any): Machine Working  
properly

Engineer's Name: \_\_\_\_\_

Engineer's Signature: \_\_\_\_\_

No Payment Done - Warranty/Rental

Customer's Signature & Seal \_\_\_\_\_

Sunit  
27/7/21

5

Date:27-07-2021 11:25  
Calibration OK

Date:27-07-2021 11:27  
Calibration OK

Date:27-07-2021 11:29  
Calibration OK

*Sund*  
*27/7/21*