

# TESTNCAL LABORATORY

(A DIVISION OF TESTING & CALIBRATION INDIA (OPC) PVT. LTD)

Plot No-113, Pocket-A2, Sector-17, Dwarka, New Delhi-110075 (India)

Mob. +91-8882056269, 9990730550, E-mail-testncal@gmail.com; Website - www.tnclab.com



## Calibration Certificate

Certificate/ULR No.: CC205522000000894F

Page 1 of 1

Customer Details		Instrument Receipt Date	1-Jan-2022
M/S SHREE PATH LAB First Floor, Ratan Ultrasound Centre Opposite Government Girls School Circular Road, Rewari Haryana Pin Code : 123401		SRF No.	TNC1/JC/21-22/130
		Date of Calibration	1-Jan-2022
		Next Calibration Date	31-Dec-2022
		Certificate Issue Date	3-Jan-2022
A) Identification On UUC (Unit Under Calibration)			
Instrument Name	Centrifuge	Make / Model No.	---/---
Range	0 to 3500 rpm	Instrument Condition	SATISFACTORY
Least Count	10 rpm	Location of Calibration	At Site
Serial No.	---	Location	Shree Path Lab Rewadi
Instrument ID.	CEN-01	Parameter	Speed
ENVIRONMENTAL CONDITION	Temperature: 25±2°C, Relative Humidity: 50±10%	Calibration Reference Standard	IS: 12508
		Calibration Procedure	TNC -02-RPM 2.0

### Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Valid upto
Dig. Tachometer	TNC/RPM/C/002/19	CC233921000001239F	10/02/2022

### Calibration Results

Sl. No.	UUC Value (rpm)	Std. Value (rpm)	Error (rpm)	Uncertainty at k=2, ± %
1	100	100	0	1.65
2	500	501	-1	0.45
3	1000	1002	-2	0.19
4	2000	2003	-3	0.19
5	3500	3504	-4	0.19

- Note:
1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
  2. This report should not be reproduced except in full without our prior permission in writing.
  3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement.
  4. All our certificates are Traceable to national/International standard.
  5. This Certificate relates to the item Calibrated total liability of Testncal Laboratory will be limited to the invoiced amount only.

Calibrated By



Issued & Approved By   
(Amit Kumar)  
CEO

Testncal Laboratory  
Plot No. 113, Pocket A2, Sector-17  
Dwarka, New Delhi-110075