

TESTNCAL LABORATORY

(A DIVISION OF TESTING & CALIBRATION INDIA (OPC) PVT. LTD)



Plot No-113, Pocket-A2, Sector-17, Dwarka, New Delhi-110075 (India)

Mob. +91-8882056269, 9990730550, E-mail-testncal@gmail.com; Website - www.tnclab.com

Calibration Certificate

Certificate/ULR No.: CC205522000000895F Page 1 of 1								
Customer Details			Instrument Receipt Date		1-Jan-2022			
M/S SHREE PATH LAB			SRF No.		TNC1/JC/21-22/130			
First Floor, Ratan Ultrasound Centre Opposite Goverment Girls School Circular Road, Rewari Haryana Pin Code : 123401			Date of Calibration		1-Jan-2022			
			Next Calibration Date		31-Dec-2022			
			Certificate Issue Date		3-Jan-2022			
A) Identification On UUC (Unit Under Calibration)								
Instrument Name	Micropipette		Make / Model No.	Dragon Lab/				
Range	100 to 1000 μl		Instrument Condition	SATISFACTORY				
Least Count	5 µl		Location of Calibration	At Laboratory				
Serial No.	YE199AL0541492		Location	Shree Path Lab Rewadi				
Instrument ID.	Pipette-01		Parameter	Volume				
ENVIRONMENTAL	Temp:- (24±2)°C, RH: - (50±10)%	Calibra	Calibration Reference Standard IS		O 8655			
CONDITION		Calibra	libration Procedure T		TNC-02-V 2.0			

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Valid upto
Weights E2 Class (1 mg to 200 g)	TNC/M/C/002/17	FL/C/M/11012021-C001	16/01/2024
Digital Weighing Balance	TNC/C/M/001/17	CC205522000000049F	02/01/2023

Calibration Results

Nominal Volume (µI)	Measured mean volume (µl)	Error (μl)	Uncertainty at k=2 (µl)			
1000	1000.89	-0.89	9.0			
500	500.24	-0.24	9.0			
100	100.37	-0.37	2.2			

Note:

- 1. The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- 3. Calibration certificate without signature are not valid and this certificate is intended of only for guidance and not for legal purpose or for advertisement
- 4. All our certificates are Traceable to national/International standard.
- 5. This Certificate relates to the item Calibrated total liability of Testnoal Laboratory will be limited to the invoiced amount only.

Calibrated By



Issued & Approved By

(Amit Kumar) CEO

Testnical Laboratory
Plot No. 113, Pocket A2, Sector-17
Dwarka, New Delhi-110075