



TESTNCAL LABORATORY

(A DIVISION OF TESTING & CALIBRATION INDIA (OPC) PVT. LTD)



Plot No-113, Pocket-A2, Sector-17, Dwarka, New Delhi-110075 (India)
Mob. +91-8882056269, 9990730550, E-mail-testncal@gmail.com; Website - www.tnclab.com

Calibration Certificate

Certificate/ULR No.: CC205522000000897F

Page 1 of 1

Customer Details

M/S SHREE PATH LAB
First Floor, Ratan Ultrasound Centre Opposite Government Girls
School Circular Road, Rewari Haryana Pin Code : 123401

Instrument Receipt Date	1-Jan-2022
SRF No.	TNC1/JC/21-22/130
Date of Calibration	1-Jan-2022
Next Calibration Date	31-Dec-2022
Certificate Issue Date	3-Jan-2022

A) Identification On UUC (Unit Under Calibration)

Instrument Name	Digital Temperature Indicator with	Make / Model No.	Mextech/---
Range	Sensor -10 to 110 °C	Instrument Condition	SATISFACTORY
Least Count	0.1°C	Location of Calibration	At Site
Serial No.	---	Location	Shree Path Lab Rewadi
Instrument ID.	DTI-01	Parameter	Thermal
ENVIRONMENTAL CONDITION	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Valid upto
4-Wire RTD with Indicator	TN/T/C/005/17	CC317221000000028F	09/02/2022

Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±)
1	20.0	20.05	0.05	0.48
2	40.0	40.15	0.15	0.48
3	60.0	60.21	0.21	0.48
4	80.0	80.24	0.24	0.48
5	110.0	110.33	0.33	0.48

- Note: -
- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
 - This report should not be reproduced except in full without our prior permission in writing.
 - Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legel purpose or for advertisement.
 - All our certificates are Traceable to national/International standard.
 - This Certificate relates to the item Calibrated.Totaliability of Testncal Laboratory. will be limited to the invoiced amount only.

Calibrated By



Issued & Approved By
CEO

Testncal Laboratory
Plot No. 113, Pocket A2, Sector-17
Dwarka, New Delhi-110075