



# TESTNCAL LABORATORY

(A DIVISION OF TESTING & CALIBRATION INDIA (OPC) PVT. LTD)



Plot No-113, Pocket-A2, Sector-17, Dwarka, New Delhi-110075 (India)

Mob. +91-8882056269, 9990730550, E-mail-testncal@gmail.com; Website - www.tnclab.com

## Calibration Certificate

Certificate/ULR No.: CC205522000000893F Page 1 of 1

<b>Customer Details</b>		Instrument Receipt Date	1-Jan-2022
<b>M/S SHREE PATH LAB</b>		SRF No.	TNC1/JC/21-22/130
First Floor, Ratan Ultrasound Centre Opposite Government Girls School Circular Road, Rewari Haryana Pin Code : 123401		Date of Calibration	1-Jan-2022
		Next Calibration Date	31-Dec-2022
		Certificate Issue Date	3-Jan-2022

<b>A) Identification On UUC (Unit Under Calibration)</b>			
Instrument Name	Incubator	Make / Model No.	---/---
Range	0 to 110 °C	Instrument Condition	SATISFACTORY
Least Count	As per Instruments	Location of Calibration	At Site
Serial No.	---	Location	Shree Path Lab Rewadi
Instrument ID.	INC-02	Parameter	°C
<b>ENVIRONMENTAL CONDITION</b>	Temperature:- 25±2 °C,Relative Humidity: - 50±10%	Calibration Reference Standard	ITS-90
		Calibration Procedure	TNC-06-T 2.0

### Details of Test Equipments Used

Name	I.D. NO./ SR NO	Certificate/ULR No.	Valid upto
4-Wire RTD with Indicator	TN/T/C/005/17	CC317221000000028F	09/02/2022

### Calibration Results

Sr.No.	UUC Value (°C)	Standard Value (°C)	Deviation (°C)	Uncertainty at k=2 at 95 % Confidence Level (±)
1	10.0	10.01	0.01	0.48
2	40.0	40.02	0.02	0.48
4	50.0	50.04	0.04	0.48
5	70.0	70.07	0.07	0.48
5	110.0	110.09	0.09	0.48

Note: -

- The Calibration results reported in this certificate are valid at the time of and the started condition of measurement.
- This report should not be reproduced except in full without our prior permission in writing.
- Calibration certificate without signature are not valid and this certificate is inteded of only for guidance and not for legei purpose or for advertisement.
- All our certificates are Traceable to national/International standard.
- This Certificate relates to the item Calibrated.Totaliability of Testncal Laboratory. will be limited to the invoiced amount only.

  
Calibrated By





Issued & Approved By  
CEO

Testncal Laboratory  
Plot No. 113, Pocket A2, Sector-17  
Dwarka, New Delhi-110075