

CALIBRATION REPORT

Centrifuges, Tabletop(Centrifuge)



Report No:	PM-CG-2021-08-6042	Date:31-8-21	Page 01 of 02
------------	--------------------	--------------	---------------

Customer Details			
Health Facility	Chirmiri CHC		
State	Chhattisgarh	District	Koriya

Medical Equipment Details			
ME Number	CGZ5KORCH1010107	Manufacturer	NA
ME Name	Centrifuges, Tabletop(Centrifuge)	Model	Na
Department	Ground Floor - HIV Room	Serial No	NA
Calibrated at	ON Site	Condition of ME	Good
Date of Calibration	04-8-21	Due for Calibration	03-8-22

Environmental Conditions			
Temperature in °C	30C	Humidity in % Rh	20%


Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/002	Digital Tachometer	Amprobe	Tach-10	17110003	23-04-2022

Performance Test						
No	Parameters	Unit	Set (rpm)	Measured (rpm)	AcceptableRange	Remarks
1	Speed	rpm	500	503	+/- 5%	Pass
			1000	972		Pass
			1500	1437		Pass
			2000	1976		Pass
			2500	2581		Pass

CALIBRATION REPORT

Centrifuges, Tabletop(Centrifuge)



Conditions of issue of report	
1. The Standards used are traceable to National standards	
2. This report should not be reproduced except in full without written approval	
Calibration Status	Passed <input checked="" type="checkbox"/> Failed <input type="checkbox"/>
Calibrated By	Authorised Signatory
Niraj Agrahari	 Swapna Smita Samal