

CALIBRATION REPORT

Incubators, Test Tube(Lab Incubator)



Report No:	PM-CG-2021-08-5973	Date: 31 Aug-2021	Page 01 of 01
------------	--------------------	-------------------	---------------

Customer Details			
Health Facility	Chirmin CHC		
State	Chhattisgarh	District	Koriya

Medical Equipment Details			
ME Number	CGZ5KORCH11010021	Manufacturer	NA
ME Name	Incubators, Test Tube(Lab Incubator)	Model	NA
Department	Ground Floor - Idc	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	04-Aug-21	Due for Calibration	03-Aug-22

Environmental Conditions			
Temperature in °C	30°C	Humidity in % Rh	51%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/0080	Digital Thermometer	METRAVI	DMT-900	171112537	23-04-22

Performance Test						
No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	Temperature Display	°C	37	36.9	36.5 to 37.5	Pass

Conditions of issue of report	
1. The Standards used are traceable to National standards.	
2. This report should not be reproduced except in full without written approval.	
Calibration Status	Passed <input checked="" type="checkbox"/> Failed <input type="checkbox"/>
Calibrated By	Authorised Signatory
Niraj Agrahari	Swapna Sudip Samal

