



COMMITTED TO THE
CUSTOMER SINCE - 1996

VAIDYANATHESHWARA INSTRUMENTS



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer

M/s.: Focus Diagnostics
72nd Cross Road, Rajajinagar,
Bangalore-560 010.

Page 1 of 2

2 Customer Reference

2.1 ULR No. :
2.2 Format No. : CC247322300000760F
2.3 SRF.No : VI-FRM-TH-004
2.4 DC No / Date : 6629
2.5 Certificate No. : --- / ---
2.6 Date of Issue : VI/21-22/6629-01
: 27-01-2022

3 Details Of Device Under Calibration(DUC),

3.1 Nomenclature : DIGITAL THERMOMETER
3.2 Make / Model : Mextech / PM-10
3.3 SI.No : M002281
3.4 DUC Condition : Satisfactory
3.5 Calibration Ref. Std. : SOP-38-04
3.6 No.of Pages : 2
3.7 Calibration Date : 27-01-2022
3.8 Calibration Due : 26-01-2023
3.9 Calibration done at : VI Thermal Lab
3.10 Discipline : Thermal

4 Environmental Condition

Temperature 25.2 °C Humidity 51 %Rh

5 Standards Used for Temperature Calibration

Sl. No.	Nomenclature	Make / Model	Sl. No.	Certificate. No.	Validity
1	Precision Thermometer With RTD Sensor	Lutron / Tm-917 / PT-100	Q601311	VI/21-22/INT-TH-55	27-08-2022
2	Dry Block Calibrator is used as source.				

6 Note:

- The Calibration Certificate relates only to the above DUC
- Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
- Corrections/erasing, invalidate the Calibration Certificate.
- Calibration of the DUC are traceable to National standards/International Standards
- Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By

Vivek W

(Calibration Engineer)

Checked By

Umesh.D

(Lab In-Charge)





COMMITTED TO THE CUSTOMER SINCE - 1996

VAIDYANATHESHWARA INSTRUMENTS



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No. VI/21-22/6629-01

Page: 2 of 2

Results:

Range : -50 to 110 °C

Resolution:- 0.1 °C

Sl No	Set Point	STD Reading (°C)	DUC Reading (°C)	Error Observed (°C)	Measurement Uncertainty ± (°C)	k Factor
1	-40	-40.14	-40.3	-0.16	0.18	2.0
2	0	0.03	-0.1	-0.13	0.18	2.0
3	30	29.97	29.8	-0.17	0.18	2.0
4	50	49.91	49.6	-0.31	0.18	2.0
5	100	99.88	99.5	-0.38	0.18	2.0

Conclusion/Remarks:-

1 Measurement uncertainty reported is at 95.45 % confidence level with k=2.

Calibrated By

Vivek W

(Calibration Engineer)

Checked By

Umesh.D

(Lab In-Charge)





COMMITTED TO THE
CUSTOMER SINCE - 1996

Vaidyanatheshwara INSTRUMENTS



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Page 1 of 2

1 Name and Address of the Customer

M/s.: Focus Diagnostics
72nd Cross Road, Rajajinagar,
Bangalore-560 010.

2 Customer Reference

2.1 ULR No. : CC247322300000761F
2.2 Format No. : VI-FRM-TH-004
2.3 SRF.No : 6629
2.4 DC No / Date : --- / ---
2.5 Certificate No. : VI/21-22/6629-02
2.6 Date of Issue : 27-01-2022

3 Details Of Device Under Calibration(DUC).

3.1 Nomenclature : DIGITAL THERMOMETER
3.2 Make / Model : Mextech / PM-10
3.3 SI.No : M002389
3.4 DUC Condition : Satisfactory
3.5 Calibration Ref. Std. : SOP-38-04
3.6 No.of Pages : 2
3.7 Calibration Date : 27-01-2022
3.8 Calibration Due : 26-01-2023
3.9 Calibration done at : VI Thermal Lab
3.10 Discipline : Thermal

4 Environmental Condition

Temperature 25.2 °C Humidity 51 %Rh

5 Standards Used for Temperature Calibration

Sl. No.	Nomenclature	Make / Model	Sl. No.	Certificate. No.	Validity
1	Precision Thermometer With RTD Sensor	Lutron / Tm-917 / PT-100	Q601311	VI/21-22/INT-TH-55	27-08-2022
2	Dry Block Calibrator is used as source.				

6 Note:

- The Calibration Certificate relates only to the above DUC
- Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
- Corrections/erasing, invalidate the Calibration Certificate.
- Calibration of the DUC are traceable to National standards/International Standards
- Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- Results Reported are valid at the time of and under the stated conditions of measurements.
- The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By

Vivek W

(Calibration Engineer)

Checked By

Umesh.D

(Lab In-Charge)



This report refers only to the item/gauge submitted and may not be reproduced except in full without written permission from Vaidyanatheshwara Instruments Bangalore - 560096.



Vaidyanatheshwara Instruments



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No. VI/21-22/6629-02

Page: 2 of 2

Results:

Range : -50 to 110 °C

Resolution:- 0.1 °C

SI No	Set Point	STD Reading (°C)	DUC Reading (°C)	Error Observed (°C)	Measurement Uncertainty ± (°C)	k Factor
1	-40	-39.86	-40.2	-0.34	0.18	2.0
2	0	-0.06	-0.1	-0.04	0.18	2.0
3	30	30.04	29.8	-0.24	0.18	2.0
4	50	50.07	49.7	-0.37	0.18	2.0
5	100	100.12	99.6	-0.52	0.18	2.0

Conclusion/Remarks:-

1 Measurement uncertainty reported is at 95.45 % confidence level with k=2.

Calibrated By

Vivek W

(Calibration Engineer)

Checked By

Umesh.D

(Lab In-Charge)

