

Calibration Certificate



Following voltages have been checked & adjusted on Blood Cell Counter Swelab Alfa (Sr. No. – 23460) Sanjivani Diagnostics, 2085, Near Delite Cinema, Behind Asaf Ali Road, Petrol Pump, Delhi-110002.

- System Voltages are in the specified range –Checked by Digital Multimeter are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Main Power Supply	219 VAC	220 +/- 5.0 VAC
2	24 Voltage Supply	24.1 VDC	24.0 +/- 1.0 VDC
3	12 Voltage Supply	12.1 VDC	12.0 +/- 0.5 VDC
4	HGB LED Voltage	4.52 VDC	4.50 +/- 3.0 VDC

- Checked the Gain and Offset of Analog Board and found OK.
➤ All instrument thru' the Service Program and found OK.
➤ Checked the Pumps & Valves and found OK.
➤ Calibrate the instrument and Run control. Control results were OK
➤ Next Calibration will be performed on 04 /04/2023.

For AAMX Biomedicals,



Nitin Choudhary
Zonal Manager

Date : 05/04/2022

INSTALLATION QUALIFICATION

The Installation for Swelab Alfa (Boule) Sr.No. 23460 carried at Sanjivani Diagnostics, 2085, Near Delite Cinema, Behind Asaf Ali Road, Petrol Pump, Delhi-110002


Checked all the Mechanisms through the Software like shear valve Movement for different position.

Checked the system performance for different Parameters .

Calibration & Control for all the Assays was satisfactory.

Carried out the Precision Check for above parameters which was also good.

For AAMX Biomedicals,



*Nitin Choudhary
Delhi*

Date : 05/04/2022

AAMXBIOMEDICALS



Shop No. 01 Shiv Mandir Tara Nagar District Sonapat-131001, State - Haryana, India
Email: aamxbiomedicals@rediffmail.com, Mob: +91 9813151502

PERFORMANCE QUALIFICATION

Carried out the following Checks for the Performance of the System Swelab Alfa (Boule) Sr. No. 23460

Run the Normal, Low, High controls (Boule)

Quality Checks for which the performance is OK & the Results are Satisfactory, within the Specified range.

Precision Checks: Carried out the Precision checks for different Parameters for which Precision was found to be satisfactory.

For AAMX Biomedicals,



Nitin Choudhary
Zonal Manager
Delhi

Date : 05/04/2022

OPERATIONAL QUALIFICATIONS


For Operational Qualifications carried out the following checks on the Swelab Alfa (Boule) Sr. No.-23460

Checked the overall System Mechanism found performance is ok.

Software Checks: Carried out the check for Calibration, Control, Results, Data & the performance is ok.

Checked all the Operational Qualifications & found the functioning is Satisfactory.

For AAMX Biomedicals,



Nitin Choudhary
Zonal Manager
Delhi

Date : 05/04/2022

INSTALLATION CERTIFICATE

This is to certify that the Swelab Alfa (Boule) Blood Cell Counter Instrument Sr.No.: 23460 has been successfully Installed and Commissioned at Sanjivani Diagnostics, 2085, Near Delite Cinema, Behind Asaf Ali Road, Petrol Pump, Delhi-110002 as per the Installation Procedure & Checklist.

For AAMX Biomedicals


For Sanjivani Diagnostics

Name : Nitin Choudhary

Name :

Designation: Zonal Manager

Designation:

Signature: 

Signature:



Date: 05/04/2022

Date: