



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Sanjivani Diagnostics Shop-2085, Ganjmer Khan, Near Delite Cinema, Opposite Asaf Ali Road, Delhi-110002	Service request No. & date	A/02/19/05/22&19/05/2022
		ULR no.	CC223922000007784F
		Cert. No.	AT22000007784
		Date of Receipt of DUC	19/05/2022
		Date of calibration	19/05/2022
		Date of issue	20/05/2022
		Suggested due date	18/05/2023

Instrument Details			
Instrument name	Micro Pipette	Location	-----
Make / Model No.	Erba	Accuracy	-----
Range / Size	0 to 50 µl	Visual Inspection	OK
Least Count	As per range	Calibration Performed at	Lab
I.D. No. / Sr. no.	----- / DY02645		

Detail of reference standards & Major equipments used			
Equipment Name	A Set of weights	Digital Weighing Balance	
Make	Weightronics	AND	
Model / SR No.	-----	GH-252/ 78002	
Certificate No.	TYCON/WB/12/274	TYE/WB/02/2022/022	
Calibration Validity	25/12/2023	20/02/2023	
Calibration by	TYCON ENGINEERING	Tycon Engineering	

Environmental Condition	Temperature	(23± 3) °C	Calibration Reference	IS:4162 (1985)
	Relative Humidity	(50±20) %	Work Instruction	AT/WI-20M

Calibration Results			
Serial No.	Nominal Capacity in (µl)	Measured Capacity in (µl)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	20.0	19.94	±0.21 µl
02.	30.0	29.87	±0.21 µl
03.	50.0	49.80	±0.21 µl

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
(PRASHANT)
Form No.- QF-47



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

--End of report--