



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Sanjivani Diagnostics Shop-2085, Ganjmer Khan, Near Delite Cinema, Opposite Asaf Ali Road, Delhi-110002	Service request No. & date	A/02/19/05/22&19/05/2022
		ULR no.	CC223922000007787F
		Cert. No.	AT22000007787
		Date of Receipt of DUC	19/05/2022
		Date of calibration	19/05/2022
		Date of issue	20/05/2022
		Suggested due date	18/05/2023

Instrument Details			
Instrument name	Centrifuge	Location	-----
Make / Model No.	REMI / R-8C	Accuracy	-----
Range / Size	0 to 4000 rpm	Visual Inspection	OK
Least Count	1 rpm	Calibration Performed at	Lab
I.D. No. / Sr. no.	----- / ZEHN-27843		

Detail of reference standards & Major equipments used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/A.C92792		
Certificate No.	QCTS/000229/10		
Calibration Validity	21/10/2022		
Calibration by	QCTS		

Environmental Condition	Temperature	20±2°C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %	Work Instruction	AT/WI-31M

Calibration Results			
Serial No.	Nominal Value (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	1000	1004	0.20%
02.	2020	2022	0.20%
03.	2830	2826	0.20%
04.	3720	3730	0.20%

Remark

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
(R.B. SAXENA)
Form No.-QF-47



Approved By
(QM/TM)
(NBERAJ YAGI)
Page No. 1 of 1

--End of report--