



# Quality Services & Laboratories

Plot No.10, Second Floor, D.S.I.D.C. Scheme - III  
Okhla Industrial Area, Phase - II, New Delhi- 110020, India,  
Tel. : +91 85953 43496  
Email: corp@qsl.co.in ; Web.: www.qslglobal.com



## CALIBRATION CERTIFICATE

Unique Lab Report No. : ULR-CC290022000002978F  
Certificate No. : QSL/DEL/CAL/22032206.02  
Date of Calibration: 22.03.2022  
Date of Receipt of Instruments: 22.03.2022

Page 1 of 1  
SRF No. : QSL/22032206  
Suggested Due Date: 22.03.2023  
Date of Issue: 24.03.2022

### Customer Name & Address:

M/S : Sachdeva Diagnostics Pvt. Ltd.,  
E-991, Saraswati Vihar,  
Delhi-110034

### Description of Instrument:

Name :	Micro Pipette	Range :	100 - 1000 µl
Make :	---	Least Count :	5 µl
S. No. :	YE16CAA0139986	Accuracy :	---
Model No. :	---	Performed at :	Lab
Location :	---	Visual Inspection :	OK
Customer ID :	---	Zero Error :	Nil

### Environmental Condition:

Temperature :	25°C ± 3 °C	Humidity :	40% to 60%
---------------	-------------	------------	------------

### Methodology of Calibration:

Reference Standard: IS: 4787	Work Instruction No. :	QSLWI/M/21
------------------------------	------------------------	------------

### Standard & Major Equipment (s) Used for calibration:

S.No.	Instrument Name	Make	S. No./Model No.	Certificate No.	Calibrated By	Valid Upto
1.	Std. Weight Box	Weightronics	---	WMCL/F/2021-06/1310	Weightronics	10.06.2023

### Calibration Result:

UUC Reading in µl	Standard Reading in µl	Error in µl	Uncertainty of Measurement:
100	100.102	-0.102	± 0.9 µl
200	200.145	-0.145	
500	500.221	-0.221	
800	800.382	-0.382	
1000	1000.460	-0.460	

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%)

### Note:

- \*The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement.
- \*This certificate cannot be reproduced except in full without our prior permission in writing.
- \*This certificate refers only to the particular item (s) calibrated.
- \*Laboratory standard are traceable to national standards.
- \*UUC: Unit Under Calibration

Prepared By 	Approved By 
Customer Service Cell (Gajendra Singh)	CEO (Mridul Kohli)