



Quality Services & Laboratories

Plot No.10, Second Floor, D.S.I.D.C. Scheme - III
Okhla Industrial Area, Phase - II, New Delhi- 110020, India,
Tel. : +91 85953 43496
Email: corp@qsl.co.in ; Web.: www.qslglobal.com

CALIBRATION CERTIFICATE

Certificate No. : QSL/DEL/CAL/2203225006.01
Date of Calibration: 22.03.2022
Date of Receipt of Instruments: 22.03.2022

Page 1 of 1
SRF No. : QSL/2203225006
Suggested Due Date: 22.03.2023
Date of Issue: 24.03.2022

Customer Name & Address:

M/S : Sachdeva Diagnostics Pvt. Ltd.,
E-991, Saraswati Vihar,
Delhi-110034

Description of Instrument:

Name :	*Centrifuge Machine	Range :	0 to 4250 RPM
Make :	Remi	Least Count :	N/A
S. No. :	----	Accuracy :	----
Model No. :	R-8C	Performed at :	Site
Location :	----	Visual Inspection :	OK
Customer ID :	CF-01	Zero Error :	Nil

Environmental Condition:

Temperature :	25°C ± 5°C	Humidity :	45 % to 75%
---------------	------------	------------	-------------

Methodology of Calibration:

Work Instruction No. : QSL/WI/ET/23

Standard & Major Equipment (s) Used for calibration:

S. No.	Instrument Name	Make	S. No. /Model No.	Certificate No.	Calibrated By	Valid Upto
1.	Digital Tachometer	----	L 142329	CC266122000000262F	MCL	21.02.2023

Calibration Result:

UUC Reading in RPM	Standard Reading in RPM	Error (RPM)	Uncertainty of Measurement:
500	500.9	-0.9	± 1 %
1000	1001	-1	
2000	2001	-1	
3000	3002	-2	
4000	4003	-3	

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%)

Note:

- *The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement.
- *This certificate cannot be reproduced except in full without our prior permission in writing.
- *This certificate refers only to the particular item (s) calibrated.
- *Laboratory standard are traceable to national standards.
- *UUC: Unit Under Calibration.
- *Is not in our NABL Scope Of Accreditation

Prepared By

Customer Service Cell (Gajendra Singh)

Approved By

CEO (Mridul Kohli)

All business are undertaken subject to company standard general business conditions, which are available on request.