

DR.VOHRA'S MEDILAB (651)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

DR.VOHRA'S MEDILAB (651)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : Nihon Kohden Celltac Alpha

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	5.03	4.93	0.78	0.63
@ Total Leucocytes (WBC) count	/c.mm	3800.00	4,314.00	0.10	-0.89
@ Haemoglobin	g/dL	15.90	15.68	0.68	0.67
@ MCV (Mean Corpuscular Volume)	fL	101.40	101.61	0.07	-0.01
@ MCH (Mean Corpuscular Hb)	pg	31.60	31.69	-0.09	0.06
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	31.10	30.90	0.00	-0.05
@ Platelet Count	1000/ μ L	135.00	105.95	0.32	0.46
@ Hematocrit	%	51.00	50.79	0.25	0.23
@ RDW CV%	%	13.00	12.88	-0.35	-0.51

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam

PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A,

Kohinoor Mall, Kiroi Road, Kurla (W),

Mumbai - 400070

End of Report