





Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF CENTRIFUGE

Certificate No:	22900001183	Issued On:	14/03/2022
ULR No:	CC214622900001183F	Calibration Date:	09/03/2022
Job Identification No:	ICS/C/AS/03/1183	Next Calibration Date:	08/03/2023
Ref. No:	SRF- Dated: 05/03/2022		

CALIBRATED FOR:	Dr. Vohra's Medilab, DSS 259, Sector-20, Panchkula - 134117 (Haryana)
------------------------	--

EQUIPMENT DESCRIPTION

Name	RPM Measurement for Centrifuge Machine, With Speed Controlling Knob		
Make/Model	---	Visual Inspection	Ok
Serial No.	---	I.D. No.	CF-01
Range	Up to 3750 RPM	Least Count	---
Location	Lab	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	LINE SEIKI	FCRI/EQL/21-22/036	FCRI, Palakkad	27/04/2021	27/04/2023

USED EQUIPMENT DETAILS: Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-RPM01 (Measurement Method)

RESULTS: Mechanical Calibration MEASUREMENT: (in RPM)

CENTRIFUGE' KNOB POSITION	CORRECTED VALUE OF TACHOMETER READING (AVG.)	Uncertainty at approx 95% C.L and coverage factor k=2
Position # 1	2758.1	± 4
Position # 2	2951.5	± 4
Position # 3	3256.4	± 5
Position # 4	3557.7	± 6
Position # 5	3754.9	± 6

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 3. Total liability of our Organisation is limited to the invoiced amount.
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Satish
Checked by

RAJESH KAPOOR
Authorised Signatory
Liaison Head