

TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF ANALYTICAL, PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF & HISTOPATHOLOGY LABORATORY EQUIPMENT.

A MSME Registered Unit, NSIC-CRISIL Rated SE 3C
A NABL Accredited Laboratory

SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

Tel. : 022-28927149
Email : info@technical-resources.in
technicalresources@gmail.com
Handset : 9819103181
Webpage : www.technical-resources.in



Calibration Certificate

Calibration field:	Mechanical-RPM	NABL Accreditation no:	CC-2112
Accreditation valid upto:	11.04.2022	Certificate no:	CC211221000000264
Requisition no:	Oct2021-139	Date of calibration:	16.10.2021
Calibration done at	At Site	Agreed Next Calib On:	15.10.2022
		Page no:	1 of 2

Customer details:		Reference used:	
Name:	Minu Clinical Laboratory	Master	Tachometer DT-2001B
Address	Sunflower Building ,Shop no.7/8 Eksar Road, Borivali(West) Mumbai-400091	Master Ser. No.	O33442 ID No. RMOO2
		Traceable Certificate	M-210720-7-6
		Valid upto	24.07.2022

Details of DUC:		Condition of item:	
Equipment:	Centrifuge		Good
Make:	Labtech	Range RPM	Not Specified
Model:	Not specified	Accuracy	Not Specified
Serial No.	MCL-EQP-07	Resolution (RPM)	Not Specified
ID No.	Not specified		
Location	Lab		

Ambient Temp(°C):	25+/- 2	Humidity Rh(%):	50 +/-20
-------------------	---------	-----------------	----------

Remark:

1. Calibration procedure used: LAB-SOP/08
2. The Reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $K=2$, providing a level of confidence of approx. 95%.
3. The observations reported represent values at the time of measurement. They donot convey any long term stability information.
4. The standards maintained are traceable to National / International Standards through accredited laboratories.
5. Sticker affixed on the unit

Note:

1. This certificate refer only to particular instrument submitted for calibration.
2. This certificate shall not be reproduced.
3. Technical Resources shall not be responsible for any change in calibration date and performance specification on account of malfunctioning of the instrument due to damage caused to it after handing over it to the customer.

Authorised signature
Sr Technical Manager
N Srinivas Rao



TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF ANALYTICAL, PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF & HISTOPATHOLOGY LABORATORY EQUIPMENT.

A MSME Registered Unit, NSIC-CRISIL Rated SE 3C
A NABL Accredited Laboratory

Tel. : 022-28927149
Email : info@technical-resources.in
technicalresources@gmail.com
Handset : 9819103181
Webpage : www.technical-resources.in



SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

Calibration data:

Equipment:	Centrifuge	Certificate no:	CC211221000000264
Serial no:	MCL-EQP-07	Date of calibration	16.10.2021
ID no.	Not specified	Page	2 of 2

RUN DETAILS

Set RPM Display	Avg. Reading on UUC Display RPM	Avg. Reading On Ref. RPM	Deviation in RPM (+/-)	Expanded Uncertainty in RPM (+/-)
SP-1	NA	2354.0	NA	25.0

END OF RESULT

20.10.2021
Certificate Issue Date:

Dilip
Checked By

Authorised signatory
Sr Technical Manager
N Srinivas Rao



F7.8.4

TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF ANALYTICAL, PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF & HISTOPATHOLOGY LABORATORY EQUIPMENT.

A MSME Registered Unit, NSIC-CRISIL Rated SE 3C
A NABL Accredited Laboratory

Tel. : 022-28927149
Email : info@technical-resources.in
technicalresources@gmail.com
Handset : 9819103181
Webpage : www.technical-resources.in



SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

Calibration Certificate

Calibration field:	Mechanical-RPM	NABL Accreditation no:	CC-2112
Accreditation valid upto:	11.04.2022	Certificate no:	CC211221000000265
Requisition no:	Oct2021-139	Date of calibration:	16.10.2021
Calibration done at	At Site	Agreed Next Calib On:	15.10.2022
		Page no:	1 of 2

Customer details:		Reference used:	
Name:	Minu Clinical Labouratory	Master	Tachometer DT-2001B
Address	Sunflower Building ,Shop no.7/8 Eksar Road, Borivali(West) Mumbai-400091	Master Ser. No.	O33442 ID No. RMOO2
		Traceable Certificate	M-210720-7-6
		Valid upto	24.07.2022

Details of DUC:		Condition of item:	
Equipment:	Centrifuge		Good
Make:	Remi	Range RPM	Not Specified
Model:	R8C	Accuracy	Not Specified
Serial No.	LELC-3515	Resolution (RPM)	Not specified
ID No.	Not specified		
Location	Lab		

Ambient Temp(°C):	25 +/- 2	Humidity Rh(%):	50 +/-20
-------------------	----------	-----------------	----------

Remark:

1. Calibration procedure used: LAB-SOP/08
2. The Reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $K=2$, providing a level of confidence of approx. 95%.
3. The observations reported represent values at the time of measurement. They donot convey any long term stability information.
4. The standards maintained are traceable to National / International Standards through accredited laboratories.
5. Sticker affixed on the unit

Note:

1. This certificate refer only to particular instrument submitted for calibration.
2. This certificate shall not be reproduced.
3. Technical Resources shall not be responsible for any change in calibration date and performance specification on account of malfunctioning of the instrument due to damage caused to it after handing over it to the customer.

Authorised signatory
Sr Technical Manager
N Srinivas Rao



TECHNICAL RESOURCES

SALES, REPAIRS, AMC & FRANCHISE SERVICES OF ANALYTICAL, PRECISION, BIO-MEDICAL, BLOOD BANKING, IVF & HISTOPATHOLOGY LABORATORY EQUIPMENT.

A MSME Registered Unit, NSIC-CRISIL Rated SE 3C
A NABL Accredited Laboratory

Tel. : 022-28927149
Email : info@technical-resources.in
technicalresources@gmail.com
Handset : 9819103181
Webpage : www.technical-resources.in



SERVICE CENTRE : OFFICE NO. 2, TONES PARADISE, OPP. NEW PEDESTRAIN BRIDGE, NAVAGAON, DAHISAR (W), MUMBAI - 400 068.

Calibration data:

Equipment:	Centrifuge	Certificate no:	CC211221000000265
Serial no:	LELC-3515	Date of calibration	16.10.2021
ID no.	Not specified	Page	2 of 2

RUN DETAILS

Set RPM Display	Avg. Reading on UUC Display RPM	Avg. Reading On Ref. RPM	Deviation in RPM (+/-)	Expanded Uncertainty in RPM (+/-)
SP-1	NA	1132.6	NA	25.0
SP-2	NA	2065.6	NA	25.0
SP-3	NA	3043.4	NA	25.0

END OF RESULT

20.10.2021
Certificate Issue Date:


Checked By

Authorised signature
Sr Technical Manager
N Srinivas Rao



F7.8.4