



RYB ENTERPRISES

Calibration & Validation Services



AN ISO : 9001:2015 Certified Company
Certified No. : DI-17021401

E-mail : rybhyderabad@gmail.com
Contact Number : +91 9963714224
+91 8867971303

CALIBRATION CERTIFICATE

For : M/s. MOTHER AND CHILD HEALTH
Integrated Counselling & Testing Centre (PPTCT LABORATORY)
Banswada Road, Banswada, Kamareddy - 503 187 Telangana.

Calibration Report No. : RYB/22/05/S007/003	Instrument ID/Sr No. : MCH/PPTCTL/CEN/01
Nomenclature : Centrifuge	SRF No. : S007
Department/Block : PPTCT Laboratory	SRF Date : 09.05.2022
Location : PPTCT LAB	Calibrated On : 15.05.2022
Make & Model : Remi	Calibration Due On : 14.05.2023
Range : 0 to 3500 RPM	Certificate Issue Date : 20.05.2022
Resolution : -NA-	Equipment Name : Centrifuge
Specified Accuracy (±) : 1% FSD	Equipment ID/Sr No. : -NA-
Environment Condition Limit	Calibration SOP NO. : WI/04
Temperature : (23±5)°C	Relative Humidity : (60±15)% rH

Sr.No	Input / Cal. Point	CALIBRATION RESULTS			Remarks	
		Master Reading	UUC Reading	Deviation		
				Observed		Acceptance Criteria (±)
Units of Measurement :		RPM				
1	500.0	498.6	500	1.4	35 RPM OK	
2	1000.0	999.1	1000	0.9		
3	2000.0	2003	2000	-3		
4	3000.0	3007	3000	-7		
5	3500.0	3492	3500	8		


CALIBRATION STANDARDS USED :

Name & Make	Range	Serial No.	Certificate No	Calibration Valid Up To
Digital Tachometer & Metravi	2 to 99999 RPM	170603433	L21101201-02	11.10.2022

REMARKS :

- The Centrifuge MCH/PPTCTL/CEN/01 is calibrated within the Accuracy.
- This report refers only to the particular item Calibrated at Site.
- The Calibration results reported in this certificate are valid at the time of & under the stated condition of environment.
- The Calibration standard used are traceable to National / International Standards.
- 'UUC' Unit under calibration, 'SRF' Service Request Form, 'SOP' Standard Operating Procedure.
- Calibration is done for points selected in full scale as per the user requirement.
- 1 gram is equal to 1000 microliters.

Note : This report should not be reproduced, without the prior permission of RYB.

 Calibrated By: RYB ENTERPRISES	Checked By: MOTHER AND CHILD HEALTH
---	-------------------------------------