

R Y B ENTERPRISES

Calibration & Validation Services



Certified No.: DI-17021401







E-mail: rybhyderabad@gmail.com Contact Number: +91 9963714224

+91 8867971303

CALIBRATION CERTIFICATE

For: M/s. MOTHER AND CHILD HEALTH

Integrated Counselling & Testing Centre (PPTCT LABORATORY)

Banswada Road, Banswada, Kamareddy - 503 187 Telangana.

Calibration Report No. : RYB/22/05/S007/004 Instrument ID/Sr No. : MCH/PPTCTL/DR/01

Nomenclature : Domestic Refrigerator Department/Block

: PPTCT Laboratory

: PPTCT LAB

Make & Model

Location

: LG & GL-B201RPZD/2022 : 2 to 8 °C

Range Resolution : -NA-

Specified Accuracy (±) : 1% FSD

Environment Condition Limit

Temperature

: (23±5)°C

SRF No. : S007

SRF Date : 09.05.2022 Calibrated On : 15.05.2022

Calibration Due On : 14.05.2023

Certificate Issue Date : 20.05.2022

Equipment Name : Domestic Refrigerator Equipment ID/Sr No. : MEZ66389410

Calibration SOP NO.

: WI/06

Relative Humidity

: (60±15)% rH

Sr.No	Input / Cal. Point	CALIBRATION RESULTS				
		Master Reading	UUC Reading	Deviation		Remarks
		waster reading	OOC Reading	Observed	Observed Acceptance	
	Units of Measurement :		°C		Criteria (±)	
1	5.0	5.041	5.0	-0.041	0.06 °C	OK

CALIBRATION STANDARDS USED:

Name & Make	Range	Serial No.	Accuracy	Certificate No	Calibration Valid Up To
PT-100(PRT) Sensor With Indicator & Yudian	-50 to 600°C	068518/ 20C107218	± 0.5 °C	L21101201-03	11.10.2022

REMARKS:

- The Domestic Refrigerator MCH/PPTCTL/DR/01 is calibrated within the Accuracy.
- This report refers only to the particular item Calibrated at Site.
- The Calibration results reported in this certificate are valid at the time of & under the stated condition of environment.
- The Calibration standard used are traceable to National / International Standards.
- `UUC' Unit under calibration, 'SRF' Service Request Form, 'SOP' Standard Operating Procedure.
- · Calibration is done for points selected in full scale as per the user requirement.
- 1 gram is equal to 1000 microliters.

Note: This report should not be reproduced, without the prior permission of RYB.

Calibrated By: RYB ENTERPRISES Checked By: MOTHER AND CHILD HEALTH