



TECHNICAL CALIBRATION CERTIFICATE

- CUSTOMER NAME : **Lahoti Medicare Pvt. Ltd.**
 - MODEL: **EM-200** SR.NO ; **B2001775-N** DATE; 26/May/2022
 - INSPECTED THE SITE FOR VIBRATION FREE PLATFORM, DUST FREE ENVIRONMENT , AWAY FROM DIRECT SUNLIGHT **YES**
 - VOLTAGE OBTAIN BETWEEN PHASE & NUTRAL : **232V**. NEUTRAL & GROUND: **01 V** AND PHASE & GROUND: **233V**.
 - INSTALLED CVT/ STABILIZER/UPS OF: **LIBERT**
 - CHECKED ALL CONNECTIONS INSIDE THE INSTRUMENT AND FOUND OK: **YES**
 - FULLY AUTOMATIC BIOCHEMISTRY ANALYZER PARAMETER CHECK (EM-360);
1. GAIN OF WAVELENGTH FILTERS :
340nm-700 405nm-530 505nm-527
546nm-511 578nm-529 600nm-456
660nm-466 700nm-416
 2. PRESSURE INDICATION : **1.05**
 3. CUVETTE BLANK VALUES WITH IN RANGE : **YES**
 4. RCT TEMPRATURE : **37.0 C (37.0 C+ 0.5)**
 5. RGT TEMPRATURE : **7.0 C (8.0 C+ 4)**
 6. LAMP VOLTAGE : **11.89V**.

NEXT CALIBRATION DUE DATE:- 25/May/2023

NAME OF THE ENGINEER: **Sunil Kumar**