



Date : 26 May'22

**Lahoti Medical Pvt. Ltd.,
Indore**

Dear Sir,
As desired by you please find the Calibration report.

1. Year of manufacturing - 2000 EasyLyte Sr. No. 45423
2. Date of Calibration - 13 May'22

Test Voltages / Test Temperature -

| Test Voltage | Actual Voltage | OK/not OK |
|------------------------|----------------|-----------|
| Logic 5 V | 5.06V | OK |
| Line voltage | 230V | OK |
| Logic 12V | 12V | OK |
| Calibration Value (Na) | 59.45 | OK |
| Calibration Value (K) | 59.71 | OK |
| Calibration Value (Cl) | 47.20 | OK |
| Electrode Values(Na) | 131 | OK |
| Electrode Value(K) | 85 | OK |
| Electrode Value(Cl) | 57 | OK |
| Pump calibration | | OK |
| Fluid Test | | OK |

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,

For **TRANSASIA BIO-MEDICALS LTD.**