Ortho Clinical
Diagnostic
Unit/Plot No - 800, ILD Trade Centre,

TO WHOM SO EVER IT MAY CONCERN

This is to certify that the Preventive Maintenance & Calibration was performed on date 10-March-2022 for the machine, VITROS 7600 Integrated Diagnostic Analyzer Installed at ESIC Medical College & Hospital, Faridabad bearing

Serial No. J76000439

The system's calibration includes checking the reproducibility performance of the instrument as per the guidelines provided by the manufacturer.

Validity: 6 Months.

Ankush Verma Service engineer Ortho Clinical Diagnostics



CALIBRATION REPORT

Instrument : VITROS 7600

Serial Number : <u>J76000439</u>

Customer : ESIC Medical College & Hospital

Department :

LABORATORY Calibration Date :

10-Mar-2022 Calibration Validity

6Months

Ortho Clinical Diagnostic Unit/Plot No - 810, ILD Trade Centre,

1. Well Wash

	Specifications	Observed Volume
Aspirated	2.0 ml – 3.0 ml	2.9 ml
Dispense	7.0 ml ml	7.0 ml

2. Signal Reagent

	Specifications	Observed Volume
Pump A	9.9 ml - 10.2 ml	10.0 ml
Pump B	9.9 ml - 10.2 ml	10.0 ml

3 Luminometer

	Specification s	Observed Volume
Luminometer Reference System	16,50,000 – 23,50,000	2027892
Incubator Reference System	16,50,000 – 23,50,000	2126880

Ortho Clinical Diagnostic Unit/Plot No - 810, ILD Trade Centre, Sector - 47, Sohna Road, Gurgaon - 122 001 T: +91 124 4014 832

SECTION/TEST

RESULT

SAMPLE METERING

PASS /FAIL a. Leak Test

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

PASS /FAIL b. Heteresis Test

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

REFERENCE METERING

a. Leak Test PASS /FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

b. <u>Heteresis Test</u> PASS /FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

Ortho Clinical Diagnostic

Unit/Plot No - 810, ILD Trade Centre, Sector - 47, Sohna Road, Gurgaon - 122 001 T: +91 124 4014 832

$\frac{\textbf{IMMUNO WASH METERING}}{a. \quad \underline{Leak \ Test}} \quad : \quad$

a. <u>Leak Test</u> : PASS /FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

b. <u>Heteresis Test</u> : PASS/FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

INCUBATOR

a. Pad Reflectance Test: PASS /FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

ELECTROMETER

a. <u>Basic Mode Verification Test:</u> PASS /FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

REFLECTOMETER

Static Test : PASS /FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

Interleaved Static Test: PASS /FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

<u>Dynamic Test</u> : PASS /FAIL

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS

OPTICS: CORRECTION FACTOR

SPECIFICATIONS	OBSERVATION
Automatic - set by software	PASS