

	1	CALIB	RATIO	N CE	RTIFI	CATE		age No:1 of 1	
CERTIFICATE NO:	SBS/CL	CALIBRATION SBS/CL/22-23/03284			THERMAL			18/36 1947 OF T	
Issue Date SRF No & Date			30-04-2022						
Receipt Date				SBS/SRF/22-23/0048-001 & 29-04-2022					
Calibration Date				29-04-2022					
Calibration Due				29-04-202				Same A A	
Customer Name & Address				28-04-2023					
ANJU DAIGNOSTIC	CENTRE.								
NO 3,RAJAMANGAL	AM, 1ST MAIN	N ROAD VILLI	VAKKAM.CH	ENNAL-49					
	- 1		ls of Device			DUC)			
Description :	TEMPER SENSOR	TEMPERATRURE INDICA SENSOR OF REFRIGERA		H Make & Model			: SAMSUNG & RR23A2F2/Y9U		
Range :	2 to 8°C			Sr. No			NA		
Least Count :	1	°C		Asset N	0				
DUC Condition	Satisfac	tory		Location			LAB		
		Environmenta	al Conditions		Alan and a second	cedure Details	E Co		
Environmental Details Temperatu		Temperatur			Relative Humidity		59% Rh		
Calibratia		SBS/CP/TH	SBS/CP/TH/04		Calibration done at		ONSITE		
	-2-17		Reference	Standard	s Details			4 -4 -1 15	
S.No Description		Make/ SI No:		Certificate No		Validity			
1 Digital mutimeter			GW Instek	GW Instek / GES 817082		HT/CC/210702-18/001		05/07/2022	
2 4 Wire RTD Sensor			TempsenDevices/TD27405			HT/CC/210	HT/CC/210702-18/004		
CALIBRATION RES	ULTS						THE RESERVE AND ADDRESS OF THE PERSON OF THE		

S.No.	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	°C	°C	°C	°C	
1	2	1.83	-0.17	0.65	
2	3	2.91	-0.09	0.65	
3	4	3.74	-0.26	0.65	
4	6	5.96	-0.04	0.65	
5	8	7.78	-0.22	0.65	

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 7. Temperature Scale:International Temperature Scale,1990(ITS-90)

Calibrated By

(Calibration Engineer)

M.BALAJI.





Chief Executive

**Authorised Signatory**