

**RAPID LABORATORY (607)  
Outlier And Analyte Summary Report**

**Outlier Details For Cycle No 220104 and Sample No 01**

**Report Date : 23/05/2022**

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**RAPID LABORATORY (607)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220104 and Sample No 01**

**Report Date : 23/05/2022**

**Instrument : Erba EM 200**


Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	76.40	70.95	0.92	0.99
@ Urea	mg/dl	23.00	22.52	0.68	0.73
@ Creatinine	mg/dl	0.81	0.80	-0.24	-0.75
@ Total Protein	g/dL	7.50	6.70	1.57	1.05
@ Albumin	g/dL	3.70	3.61	0.45	-0.17
@ Uric Acid	mg/dl	4.50	4.68	-0.14	0.38
@ ALP	U/L	45.00	54.09	-1.06	-0.48
@ Calcium	mg/dl	7.80	8.05	-0.47	-0.69
@ Cholesterol	mg/dl	145.00	142.68	0.37	0.45
@ Triglycerides	mg/dl	103.00	106.75	-0.20	-0.02
@ HDL	mg/dl	37.10	35.50	0.45	0.54
@ ALT	U/L	6.10	8.61	-0.81	-0.32
@ AST	U/L	17.10	16.87	0.20	0.46

**Instrument : HDC- LYTE**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Sodium	mmol/L	133.00	--	0.76	1.17
@ Potassium	mmol/L	3.67	--	-0.63	-0.43

Legend @ : z score  $\leq$  2.0 - Acceptable  
 ! : 2.0 < z score < 3.0 - Warning  
 X : z score  $\geq$  3 - Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	15
Not Evaluated Parameters	0
Evaluated Parameters	15
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070