

RAPID LABORATORY (607)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 02

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

RAPID LABORATORY (607)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 02


Report Date : 23/05/2022

Instrument : ERBA H 560

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	5.61	5.52	-0.04	-0.29
@ Total Leucocytes (WBC) count	/c.mm	5900.00	6,772.50	-0.03	-0.58
@ Haemoglobin	g/dL	14.60	14.40	0.00	0.01
@ MCV (Mean Corpuscular Volume)	fL	93.30	92.84	0.18	0.32
@ MCH (Mean Corpuscular Hb)	pg	25.50	25.96	-0.60	-0.22
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	27.30	28.00	-1.33	-0.66
@ Platelet Count	1000/ μ L	337.00	357.75	0.05	1.27
@ Hematocrit	%	52.40	51.75	-0.12	0.07
@ RDW CV%	%	17.00	17.98	0.01	-0.88

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam

PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A,

Kohinoor Mall, Kiroi Road, Kurla (W),

Mumbai - 400070

End of Report