

Date: 02/12/2021
Effective Date: 02/12/2021

Certificate of Calibration

Customer Name: Rapid Laboratories

Model : Automated Hematology Analyzer H560

Serial No.: K11042036144

Calibration Done Date: 02-12-2021

Next Calibration Due Date On or Before: 03/12/2022

Lab In-charge: Dr. Ish Pandhi

This is to certify that the above-mentioned product has been verified of calibration for CBC 5 parameters (WBC, RBC, HGB, MCV and PLT) according to the standard procedures provided by Erba Lachema s.r.o, Karasek.

Calibration at site performed by
Rashpal Singh Rana
ISM
Transasia Bio-Medicals Ltd
Chandigarh

Incl:
Certificate of Inspection
Assay Sheet of Hematology Calibrator (H Cal)
Printouts
Traceability Document

L-J QC Analysis

File No.: 1 Lot No.: EH2111 Level: Low
 Operator: admin Exp. Date: 2022/01/10 Run Time: 2021/12/02 11:45
 Print Time: 2021/12/02 11:49 QC Mode: Whole Blood-CBC+DIFF QC Sample ID:

Parameter	Result	Ref. Range	Unit
1 WBC	3.42	2.92-3.92	10 ³ /uL
2 Neu%	48.4	40.0-58.0	%
3 Lym%	39.2	29.0-47.0	%
4 Mon%	5.9	2.2-10.2	%
5 Eos%	6.5	1.8-11.8	%
6 Bas%	2.0	0.0-5.2	%
7 Neu#	1.66	1.28-2.08	10 ³ /uL
8 Lym#	1.34	0.90-1.70	10 ³ /uL
9 Mon#	0.20	0.07-0.35	10 ³ /uL
10 Eos#	0.22	0.08-0.38	10 ³ /uL
11 Bas#	0.07	0.00-0.18	10 ³ /uL
12 RBC	2.36	2.15-2.51	10 ⁶ /uL
13 HGB	5.9	5.4-6.2	g/dL
14 HCT	19.0	16.6-20.6	%
15 MCV	80.4	74.6-84.6	fL
16 MCH	25.0	22.5-27.5	pg
17 MCHC	31.1	28.4-34.4	g/dL
18 RDW-CV	15.3	12.8-18.8	%
19 RDW-SD	51.1	43.2-63.2	fL
20 PLT	76	49-89	10 ³ /uL
21 MPV	9.8	6.6-12.6	fL
22 PDW-SD	13.6	11.3-17.3	fL
23 PDW-CV	16.8	13.8-19.8	%
24 PCT	0.075	0.017-0.117	%
25 P-LCR	36.4	27.3-43.3	%
26 P-LCC	28	10-40	10 ³ /uL

