

ST 9283

HOLD
-115.0

HISET LOWSET ALUP HOLD °C/°F ON/OFF
MAX/MIN

-50°C TO +300°C -58°F TO +572°F

INDIAN CALIBRATION SERVICES DELHI-92

SR. NO. _____
INST. ID. DTWS-01
CAL. DATE 25-11-2021
DUE DATE 24-11-2023

CALIBRATED Ph: 011-22463000 Fax: 011-22463277
Mail: ics@icrsd.com icrsd@gmail.com

MULTI -Thermometer