

Result Reprint

Report Type : Controls 25-Jan-2022

Sr #	Lot #	Consumable	Test	Result Unit	Flag	Curve #	Result Date	Mean	SD	Interval (3SD)
1	S06212125B	ERBA NORM	GLU	99.6 mg/dl		129308	25-Jan-2022 13:02:50	95.720	4.780	81.38 - 110.06
2	S06212125B	ERBA NORM	TRIG	118.2 mg/dl		129309	25-Jan-2022 13:03:08	113.350	5.670	96.34 - 130.36
3	S06212125B	ERBA NORM	SGPTD	39.5 U/L		129310	25-Jan-2022 13:03:27	38.390	2.560	30.71 - 46.07
4	S06212125B	ERBA NORM	SGOTD	47.9 U/L		129311	25-Jan-2022 13:03:45	49.890	3.230	40.2 - 59.58
5	S06212125B	ERBA NORM	ALPU	115 U/L		129312	25-Jan-2022 13:04:03	120.390	8.520	94.83 - 145.95
6	S06212125B	ERBA NORM	BIT	1.49 mg/dl		129313	25-Jan-2022 13:04:21	1.510	0.130	1.12 - 1.9
7	S06212125B	ERBA NORM	UA	6.1 mg/dl		129316	25-Jan-2022 13:05:15	6.060	0.300	5.18 - 6.98
8	S06212125B	ERBA NORM	ALB	3.40 g/dl		129318	25-Jan-2022 13:05:51	3.570	0.180	3.03 - 4.11
9	S06212125B	ERBA NORM	HDL	44.5 mg/dl		129321	25-Jan-2022 13:06:45	42.860	2.860	34.28 - 51.44
10	S06212125B	ERBA NORM	CHOL	139 mg/dl		129328	25-Jan-2022 13:22:06	137.570	6.880	116.93 - 158.21
11	S06212125B	ERBA NORM	CA	8.4 mg/dl		129331	25-Jan-2022 13:23:00	8.090	0.400	6.89 - 9.29
12	S06212125B	ERBA NORM	PRO	6.41 g/dl		129334	25-Jan-2022 13:23:54	6.240	0.310	5.31 - 7.17
13	S06212125B	ERBA NORM	CREA	1.25 mg/dl	F,1SD	129337	25-Jan-2022 13:24:48	1.380	0.100	1.08 - 1.68
14	S06212125B	ERBA NORM	UREA	32.3 mg/dl		129339	25-Jan-2022 13:51:18	33.970	1.700	28.87 - 39.07