

Dear NABL Assessors,
As per ROTEK, the values of calibration are already specified in the below report include the details of the tachometer used for calibration
However, as stated in our undertaking letter we will be approaching an NABL accredited 17025 calibration laobaoratory for the calibration of these equipments.
Mr. Bose EK
Nirnay



Since 1993

ROTEK INSTRUMENTS

West Vengola, Perumbavoor - 683 556, Ernakulam Dist., Kerala
Email : mail@rotekinstruments.com / info@rotekinstruments.com
www.rotekinstruments.com

GST - 32AAKFR2116A1ZX

Phone : 70341 67000
85930 77000
85930 00606

CALIBRATION CERTIFICATE

Certificate no: RTK /003/2022

Dealer

M/s Life Care Diagnostics
Trivandrum

Customer:

Nirnya Laboratory
Karakkonam
Trivandrum

Date of Calibration: 18/04/2022
Calibration up to : 17/04/2023

This is to certify that the following item has been calibrated and functionally checked as per our factory norms.

INSTRUMENT: Centrifuge

Cat. No. RCH - 8
RPM: Maximum: 4000
Watts : 40
AMP S : 1.2

Sl. No. 19314

Motor : 1/20 HP Universal
Checked at : 220V AC

Tachometer Make used for calibration: TECLOCK Corporation ,Japan

CALIBRATED BY.

Sali m v

ENGINEER PRODUCTION



APPROVED BY.

Balaktishnar m v



ISO 9001:2015



EN ISO 13485:2016