



South Indian Medical System

Yoga Athipathi Building, 62/154, Rajastreet, Ganapathy, Coimbatore - 641 006

E-mail : sims.avm@gmail.com

Mobile : 9994634962, 6381504923

CALIBRATION CERTIFICATE MICRO PIPETTE

Calibration Certificate No : CAL/AKC /2022/04

Customer Name and Address:	ABIRAMI KIDNEY CARE DR.THANGAVELU HOSPITAL, ERODE.	Equipment ID:	****
Department:	LAB	Serial No:	NIL
Make:	VERTEX	Date of Calibration:	15-05-2022
Model:	NIL	Next Calibration Due:	15-05-2023

Calibration Results:

S.No	TEST VOL(ml)	Measured Value	Acceptable Range	Remarks
1	50µl	50.06	0.5ml	pass
2	50µl	49.87	0.5ml	pass
3	50µl	50.07	0.5ml	pass
4	50µl	50.04	0.5ml	pass

Remarks

1. The above equipment calibrated at site.
2. The above Tolerance level has been defined from the manufacturer's manual.
3. The above results relates only to the item calibrated.

Calibrated by:


Calibration Engineer

Approved by:


Authorized Signatory



Page 1



South Indian Medical System

Yoga Athipathi Building, 62/154, Rajastreet, Ganapathy, Coimbatore - 641 006

E-mail : sims.avm@gmail.com

Mobile : 9994634962, 6381504923

CALIBRATION CERTIFICATE MICRO PIPETTE

Calibration Certificate No : CAL/AKC /2022/05

Customer Name and Address:	ABIRAMI KIDNEY CARE DR.THANGAVELU HOSPITAL, ERODE.	Equipment ID:	****
Department:	LAB	Serial No:	NIL
Make:	VERTEX	Date of Calibration:	15-05-2022
Model:	NIL	Next Calibration Due:	15-05-2023

Calibration Results:

S.No	TEST VOL(ml)	Measured Value	Acceptable Range	Remarks
1	50µl	50.02	0.5ml	pass
2	50µl	49.92	0.5ml	pass
3	50µl	50.03	0.5ml	pass
4	50µl	50.07	0.5ml	pass

Remarks

1. The above equipment calibrated at site.
2. The above Tolerance level has been defined from the manufacturer's manual.
3. The above results relates only to the item calibrated.

Calibrated by:


Calibration Engineer

Approved by:

