

# **SOUTH INDIAN MEDICAL SYSTEM**

Yoga Athipathi Building, 62/154, Raja Street, Ganapathy, Coimbatore - 641006

E- Mail: sims.avm@gmail.com PH: 9994634962, 6381504923

# CALIBRATION CERTIFICATE MICROSCOPE

Calibration Certificate No: CAL/AKC/2022/09

Customer Name and Address:	ABIRAMI KIDNEY CARE DR.THANGAVELU HOSPITAL, ERODE.	Equipment ID:	***	
Department:	LAB	Serial No:	151185019	
Make:	LABOMED	Date of Calibration:	15-05-2022	
Model:	LX 200	Next Calibration Due:	15-05-2023	

### **Traceability Details:**

S. No	Standards Used	Make	Model	Serial No	Calibration on	Calibration Due
<u>1</u>	LUX METER	<u>HTC</u>	LX-103	<u>T119527</u>	02-05-2022	09-05-2023
<u>2</u>	ELECRTICAL SAFETY ANALYZER	RIGEL MEDICAL	RIGEL 288+	36G-1040	11-04-2022	11-04-2023

## **Equipment Condition at the time of Calibration:**

#### **Calibration Results:**

S.NO	DUC KNOP POSITION	STANDARD READING(LUX)		
1	HIGH	2800		
2	LOW	11800		

Page 1



# **SOUTH INDIAN MEDICAL SYSTEM**

Yoga Athipathi Building, 62/154, Raja Street, Ganapathy, Coimbatore - 641006

E- Mail: sims.avm@gmail.com PH: 9994634962, 6381504923

**Equipment Passed Electrical Safety Test** 

## **Remarks**

1. The above equipment calibrated at site.

- 2. The above Tolerance level has been defined from the manufacturer's manual.
- 3. The above results relates only to the item calibrated.

Calibrated by:

**Calibration Engineer** 

Approved by

**Authorized Signatory** 



# SOUTH INDIAN MEDICAL SYSTEM

Yoga Athipathi Building, 62/154, Raja Street, Ganapathy, Coimbatore - 641006

E- Mail: <u>sims.avm@gmail.com</u> PH: 9994634962, 6381504923

MICROSCOPE

ABIRAMI KIDNEY CARE CALIBRATION

ELECTRICAL SAFETY TEST RIGEL MEDICAL, RIGEL 288+

File: MICROSCOPE Date: 2022 MAY 15 Time: 12:30:22

ESA SL.NO: 36G-1040 Equipment Information...

Description: MICROSCOPE

Manufacturer: \*\*\*\*
Model: \*\*\*\*

Serial No: **151185019** 

Location: LAB

Facility: Technician ID:

Tech Time (hrs): 5min
Overall Result: PASS

Signature: Date: 15.05 2022

\*\*\*\*

Detailed Test Data. . .

Elec. Safety Standard: IEC62353

Earth Resistance Limit: Maximum 0.300 ohms

Equipment Leakage Limit: Maximum 500.0  $\mu A$  ( Class I) Applied part Lkg Limit: Maximum 5000  $\mu A$  ( Type BF)

Mains Voltage L1-L2: 228.5 V Mains Voltage L1-PE: 225.1 V Mains Voltage L2-PE: 3.4 V External Voltage: 1.2 V External milliVolts: 3.2 mV

Load Current : 1.5 A Equipment Power: 18 W

Equipment Leakage (NP): 8 μA – PASS Equipment Leakage (NP): 0.8 μA – PASS Equipment Leakage (NP): 0.9 μA - PASS

END OF RECORD

THE POOR STATE OF THE POOR STA

Approved by

Page 3