

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable † : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: Nil

Corrective Action: Nil

Reviewed by: Visharma

Dated: 23.05.22

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)


Outlier And Analyte Summary Report
 Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : Biorad D-10

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	6.60	6.94	-1.20	-1.43

Legend @ : z score \leq 2.0 - Acceptable	Total Parameters	1
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score \geq 3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
🕒 : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kuria (W),
 Mumbai - 400070



End of Report

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 02

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: Nil

Corrective Action: Nil

Reviewed by: Visharma

Dated: 23.05.22

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 02

Report Date : 23/05/2022

Instrument : Biorad D-10

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	5.60	5.74	-0.69	-1.43

Legend	@ : z score \leq 2.0 - Acceptable	Total Parameters	1
	! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
	X : z score \geq 3 - Unacceptable	Evaluated Parameters	1
	# : Not Evaluated	Outlier Parameters (X)	0
	🕒 : Delayed Result Entry	EQAS Score HbA1c	100.00 %
	* : Not considered for evaluation.		

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

Visharma

End of Report