

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit
# Other findings	-	Squamous Epithelial Cells	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: Parameter other findings not evaluated

Corrective Action: Sufficient peers are not available with EQAS provider for evaluation.

Reviewed by: Visharma

Dated: 23.05.22

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Crystal	-	Triple phosphate crystal	Triple phosphate crystal
@ Casts	-	No cast seen	No cast seen
# Other findings	-	Squamous Epithelial Cells	-

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	1
Evaluated Parameters	2
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Sediment Analysis	100.00 %

Image

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416,
Commercial Building - 1A,
Kohinoor Mall, Kirof Road, Kurla (W),
Mumbai - 400070

Visharma
23.05.22

End of Report

Certificate of Participation

Metropolis Healthcare Ltd. EQAS

Cycle 2021 – January to December

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (HLL Lifecare Ltd. Gurgaon)

Lab ID: 549

Has successfully participated in Scheme for

●M(EL)T ●IMMUNOASSAY ●Clinical Pathology



Dr. Puneet Kumar Nigam

Dr Puneet Kumar Nigam

PT Coordinator

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210512 and Sample No 01

Report Date : 30/12/2021

Analyte	Instrument	Result Value	Standard Unit
# Blood (urine)	Uriplus 600	1+	-
# Nitrite (urine)	Uriplus 600	Present	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: Blood and Nitrite was not evaluated by EQAS provider due to less peer submission.

Corrective Action:

Reviewed by:

Visharma

Dated:

30-12-21

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Result Details For Cycle No 210512 and Sample No 01

Report Date : 30/12/2021

Instrument : Uriplus 600

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5.5,6,6.5
@ Specific Gravity	-	1.025	1.015,1.020,1.025
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,Tr
# Blood (urine)	-	1+	
# Nitrite (urine)	-	Present	
@ Ketone (urine)	-	Negative	Negative
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	2
Evaluated Parameters	7
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416,
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070

Visharma

End of Report

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: Visharma

Dated: 19-07-21

SHRI MATA SHEETLA DEVI SWARNA JAYANTI DIAGNOSTIC CENTRE (549)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Instrument : Uriplus 600

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	6.6.5
@ Specific Gravity	-	1.020	1.010,1.015,1.020
@ Glucose (urine)	-	2+	+1,+2,+3
@ Protein (Urine)	-	1+	+1,+2
@ Ketone (urine)	-	Negative	Negative
@ Blood (urine)	-	Negative	Negative
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
@ Nitrite (urine)	-	Present	Present

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070

Visharma

End of Report