



Calibration Certificate

Customer Name	Shri Mata Sheetla Devi Swarna Jayanti Diagnostic Centre Hindlabs HLL Lifecare Ltd. (A Govt. of India Enterprise)
Address	Sec-39, Jharsa, Gurugram -122003
Contact Person	Dr. Vijay Kumar Sharma
Contact No.	+91-8800336556
Instrument	Urine Dry Chemistry Analyzer
Model	Uri Plus-600
Serial No.	NI900500H0348
Instrument Status	Under Warranty
Calibration Done On	27.07.2021
Next Calibration Due On	26.07.2022

For Aspen Diagnostics Pvt. Ltd.

for
Hemant Singh Bisht

Hemant Singh Bisht
Technical Executive



Rajaram Krishnamurthy

Rajaram Krishnamurthy
National Service Manager



RAPID DIAGNOSTIC GROUP OF COMPANIES

* Rapid Diagnostic Pvt. Ltd. * Aspen Diagnostics Pvt. Ltd.
 * Aspen Laboratories Pvt. Ltd. * Deep Meditech Pvt. Ltd.

B-82, G.T.K. Road Industrial Area Delhi - 110 033 (INDIA)

Ph.: 27113030, 45201100, Fax: 27413437, 45201198

WEBSITE : www.rapiddiagnostic.in



TECHNICAL SERVICE / INSTALLATION REPORT

Customer's Name and Address: Hind labs, Gurugram
Shri Mata Sheela Devi Swarna Jayanti
Diagnostic center.

Date: 27/07/2021 TSR No.: **2550**

Inst. Model: URI 600

Cont. Person: Dr VIJAY K SHARMA

Inst. S. No.: N1900500410348SE

Contact No.: 8800356516 Email: _____

Service Status: Warranty R/R A.M.C. C.M.C. Chargeable Others

Call Status: **INSTALLATION** Breakdown PM Routine Service Other

Problem Reported/ Installation Details:

* Calibration Problem,
Calibration Not okay

Instrument attended and is working satisfactorily

Job to be continued

Following parts are required to be replaced

Following parts have been replaced

Observations: * Dirty lens

* Calibration test okay
after maintenance

Sr. No.	Parts	Old Hologram	New Hologram	Price (Rs.)
1.				
2.				
3.				
4.				
5.				
Total (Rs.)				

Action Taken: * Cleaned the dirty lens

* Ran Calibration Strip
thrice

Customer's Remarks (If any): _____

Engineer's Remarks (If any): Machine Working properly

Engineer's Name: _____

No Payment Done - Warranty/Rental

Engineer's Signature: _____

Customer's Signature & Seal

Sunit
27/7/21

5

Date:27-07-2021 11:25
Calibration OK

Date:27-07-2021 11:27
Calibration OK

Date:27-07-2021 11:29
Calibration OK

Sund
27/7/21