

ASIAN TECHNOLOGY



(A HOUSE OF CALIBRATION)

CALIBRATION CERTIFICATE

Customer name	Swarna Jayanti Diagnostic Centre,	Service request No. & date	P/04/16/03/22&16/03/202
And		ULR no.	CC223922000001301F
Address Hindlabs HLL Lifecare Limited. MCG Engg. Office Building, Sector- 39, Jharsa, Gurugram-122003, Haryana, INDIA.	Cert. No.	AT/22000001301	
	39, Jharsa, Gurugram-122003,	Date of Receipt of DUC	16/03/2022
		Date of calibration	16/03/2022
		Date of issue	17/03/2022
		Suggested due date	15/03/2023

	Instrument D	etails	
Instrument name	Centrifuge RPM	S.No.	ZFHN-27772
Make /Model No	Remi/NEYA-4	Location	W 00 10 10 10 10
Range / Size	50 to 6000 rpm	Accuracy	
Least Count	10 rpm	Visual Inspection	OK
I.D. No.	HLL/HR55/GGN/CNFG/04	JALL TIL	

Detail of reference standards & Major equipments used			
Equipment Name	Digital Hand Tachometer		
Make	LT Lutron		
Model / SR No.	A.C92792		
Certificate No.	QCTS/000229/10		
Calibration Validity	21/10/2022		
Calibration by	Quality Calibration Testing Solutions		

Environmental	Temperature	20±2°C	°C Calibration Reference IS: 12508	
Condition	Relative Humidity	50±10 %	Calibration Procedure	CP-37

Calibration Results			
Serial No.	Standard value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	503.5	502	
02.	1002	1000	THE PROPERTY OF THE
03.	2010	2008	
04.	3005	3002	±0.20 %rdg
05.	5009	5005	
06.	5999	5908	

Remarks:

(1) Standard equipment use for calibration are traceable to national/international standards.
 (2) The reported expended uncertainty of measurement is stated as the standard uncertainty.

(2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factork = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Calibration Engg./TM) (MARIPAL) Form No.-QF-47



Approved By (QM/TM)
(NEERAL TYAGI)
Page No. 1 of 1