

ASIAN TECHNOLOGY

Y

C-2239

(A HOUSE OF CALIBRATION)

CALIBRATION CERTIFICATE

Customer name	M/S: Shri Mata Sheetla Devi	Service request No. & date	P/04/16/03/22&16/03/2022
And	Swarna Jayanti Diagnostic Centre,	ULR no.	CC223922000001309F
Address	Address Hindlabs HLL Lifecare Limited. MCG Engg. Office Building, Sector- 39, Jharsa, Gurugram-122003, Haryana, INDIA.	Cert. No.	AT/22000001309
		Date of Receipt of DUC	16/03/2022
		Date of calibration	16/03/2022
		Date of issue	17/03/2022
		Suggested due date	15/03/2023

Instrument Details			
Instrument name	Orbital Shaker	S.No.	ZEKS-36658
Make /Model No	Remi/RS-12R	Location	
Range / Size	08 to 180 rpm	Accuracy	(a. a. a
Least Count	5 rpm	Visual Inspection	OK
I.D. No.	HLL/HR55/GGN/MIX/01		

Equipment Name	Digital Hand Tachometer	
Make	LT Lutron	
Model / SR No.	A.C92792	
Certificate No.	QCTS/000229/10	
Calibration Validity	21/10/2022	
Calibration by	Quality Calibration Testing Solutions	

Environmental	Temperature	20±2°C	Calibration Reference	IS: 12508
Condition	Relative Humidity	50±10 %	Calibration Procedure	CP-37

Calibration Results

Serial No.	Standard value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)	
01.	41.8	40		
02.	49.2	50		
03.	98.7	100	The second secon	
04.	118.2	120	±0.20 %rdg	
05.	175.9	= 180		

Remarks:

(1) Standard equipment use for calibration are traceable to national/ international standards.

(2)The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factork = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By (Calibration Engg./TM) (MAPIPAL) Form No.-QF-47



Approved By
(QM/TM)
(NEERAL/TYAGI)
Page No. 1 of 1