



CALIBRATION CERTIFICATE

JOB ORDER NO	AIC/SRF/22-23/220536	PAGE NUMBER	01 of 02
CERTIFICATE NUMBER	AIC/22-23/220536-01	ULR NO:CC267422000002276F	Date of Issue:25-05-2022
CUSTOMER ADDRESS	M/S GOVERNMENT HOSPITAL, ODDENCHATRAM, DINDIGUL.		

DEVICE UNDER CALIBRATION DETAILS

NOMENCLATURE	MICRO PIPETTE
MAKE	MICROLIT
MODEL NO	07101874
SERIAL NUMBER	-
RANGE	10 to 100 µl
RESOLUTION	0.2 µl
LOCATION	LAB
ID NUMBER	-
DEVICE UNDER CALIBRATION RECEIVED DATE	24/May/22
CALIBRATION DATE	24/May/22
RECOMMENDED NEXT DUE DATE	23/May/23
CONDITION OF ITEM ON RECEIPT	GOOD

REFERENCE STANDARDS DETAILS

	MASTER 1	MASTER 2	MASTER 3
NOMENCLATURE	Electronic Balance		
MAKE	Radwag		
MODEL NO	AS60/220.R2		
SERIAL NUMBER	500722		
RANGE	0 to 220g		
TRACEABILITY DETAILS	TC/8471/2022		
VALIDITY	25-Feb-23		

MONITORED ENVIRONMENTAL CONDITIONS

TEMPERATURE (°C)	21.3	RELATIVE HUMIDITY (%RH)	46
CALIBRATION PROCEDURE	CP-M08-2016	STANDARD USED	ISO 8655 - 6

<p>CALIBRATED BY</p>  <p>R.UVARAJ</p> <p>CALIBRATION ENGINEER</p>	 <p>SEAL OF ARROW</p>	<p>APPROVED BY</p>  <p>A.MASILAN</p> <p>TECHNICAL MANAGER</p>
--	--	--



CC-2674

E-mail : arrowinstrument@yahoo.co.in

CERTIFICATE NUMBER	AIC/22-23/220536-01
PAGE NUMBER	02 OF 02

CALIBRATION RESULTS




DUC Readings	Standard Readings @27°C	Deviation	Systematic Error	Random Error	Expanded Uncertainty (±)
µl	µl	µl	%	%	µl
10	10.06	-0.06	-0.62	0.03	0.42
20	20.10	-0.10	-0.48	0.04	0.43
40	40.14	-0.14	-0.35	0.08	0.49
80	80.34	-0.34	-0.42	0.16	0.75
100	100.16	-0.16	-0.16	0.08	0.48

REMARKS

- 1.The above mentioned device under calibration was calibrated at Lab.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 6.Any correction / Modification in this calibration report kindly inform with in 30 days.
- 7.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
8. $V_{t2}=V_{t1}[1+g(t_2-t_1)]$ - This formula is used to compensated the thermal expansion factors at 27°C.
- 9.Calibration of the DUC are traceable to National Standards/International Standards.

MEASURED ENVIRONMENTAL CONDITIONS@LAB

Temperature (°C)	21.3	
Relative Humidity (% RH)	46	
Barometric Pressure (hPa)	966.5	
Water Temperature (°C)	20.3	
Density of water at 21.2 °c = ρw	1.139	µl/mg
Z Factor =	1.0028	µl/mg

CALIBRATED BY  R.UVARAJ CALIBRATION ENGINEER	 SEAL OF ARROW	APPROVED BY  A.MASILAN TECHNICAL MANAGER
---	---	---