

**RAPID LABORATORY (607)
Outlier And Analyte Summary Report**

Outlier Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

RAPID LABORATORY (607)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : Erba EM 200


Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	76.40	70.95	0.92	0.99
@ Urea	mg/dl	23.00	22.52	0.68	0.73
@ Creatinine	mg/dl	0.81	0.80	-0.24	-0.75
@ Total Protein	g/dL	7.50	6.70	1.57	1.05
@ Albumin	g/dL	3.70	3.61	0.45	-0.17
@ Uric Acid	mg/dl	4.50	4.68	-0.14	0.38
@ ALP	U/L	45.00	54.09	-1.06	-0.48
@ Calcium	mg/dl	7.80	8.05	-0.47	-0.69
@ Cholesterol	mg/dl	145.00	142.68	0.37	0.45
@ Triglycerides	mg/dl	103.00	106.75	-0.20	-0.02
@ HDL	mg/dl	37.10	35.50	0.45	0.54
@ ALT	U/L	6.10	8.61	-0.81	-0.32
@ AST	U/L	17.10	16.87	0.20	0.46

Instrument : HDC- LYTE

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Sodium	mmol/L	133.00	--	0.76	1.17
@ Potassium	mmol/L	3.67	--	-0.63	-0.43

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	15
Not Evaluated Parameters	0
Evaluated Parameters	15
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070