

**RAPID LABORATORY (607)**

**Outlier And Analyte Summary Report**

**Outlier Details For Cycle No 220104 and Sample No 01**

**Report Date : 23/05/2022**

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**RAPID LABORATORY (607)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220104 and Sample No 01**


**Report Date : 23/05/2022**

**Instrument : Erba EM 200**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	6.60	6.36	0.63	0.79

<b>Legend</b> @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.	Total Parameters	1
	Not Evaluated Parameters	0
	Evaluated Parameters	1
	Outlier Parameters (X)	0
	EQAS Score HbA1c	100.00 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070

End of Report