

**RAPID LABORATORY (607)
Outlier And Analyte Summary Report**

Outlier Details For Cycle No 220104 and Sample No 02

Report Date : 23/05/2022

| Analyte | Instrument | Result Value | Standard Unit | Z-Score |
|---|------------|--------------|---------------|---------|
| No Outlier | | | | |
| Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation. | | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

RAPID LABORATORY (607)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 02

Report Date : 23/05/2022

Instrument : Erba EM 200

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|-------------------|--------------|--------------|--------|---------|-------|
| @ Glucose | mg/dl | 303.00 | 288.48 | 0.75 | 0.99 |
| @ Urea | mg/dl | 21.80 | 20.79 | 0.83 | 0.73 |
| @ Creatinine | mg/dl | 2.60 | 2.82 | -1.01 | -0.75 |
| @ Bilirubin Total | mg/dl | 0.42 | 0.44 | -0.40 | 0.26 |
| @ Total Protein | g/dL | 9.00 | 8.41 | 0.79 | 1.05 |
| @ Albumin | g/dL | 5.09 | 4.96 | 0.46 | -0.17 |
| @ AST | U/L | 16.70 | 17.95 | 0.17 | 0.46 |
| @ ALP | U/L | 34.00 | 43.18 | -1.05 | -0.48 |
| @ Uric Acid | mg/dl | 4.30 | 4.01 | 0.74 | 0.38 |
| @ Calcium | mg/dl | 7.50 | 7.55 | -0.18 | -0.69 |
| @ Cholesterol | mg/dl | 138.00 | 132.33 | 0.54 | 0.45 |
| @ Triglycerides | mg/dl | 97.40 | 99.73 | -0.25 | -0.02 |
| @ HDL | mg/dl | 36.50 | 34.06 | 0.53 | 0.54 |
| @ ALT | U/L | 5.80 | 8.96 | -1.49 | -0.32 |

Instrument : HDC- LYTE

| Analyte | StandardUnit | Result Value | Mean | Z-Score | RMZ |
|-------------|--------------|--------------|------|---------|-------|
| @ Potassium | mmol/L | 3.60 | -- | -0.33 | -0.43 |

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

| | |
|--------------------------|----------|
| Total Parameters | 15 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 15 |
| Outlier Parameters (X) | 0 |
| EQAS Score Biochemistry | 100.00 % |

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070