

Date: 02/12/2021  
Effective Date: 02/12/2021

## Certificate of Calibration

Customer Name: Rapid Laboratories

Model : Automated Hematology Analyzer H560

Serial No.: K11042036144

Calibration Done Date: 02-12-2021

Next Calibration Due Date On or Before: 03/12/2022

Lab In-charge: Dr. Ish Pandhi

*This is to certify that the above-mentioned product has been verified of calibration for CBC-5 parameters (WBC, RBC, HGB, MCV and PLT) according to the standard procedures provided by Erba Lachema s.r.o, Karasek.*

Calibration at site performed by  
Rajpal Singh Rana  
FSM  
Transasia Bio-Medicals Ltd  
Chandigarh

Incl:

1. Certificate of Inspection
2. Assay Sheet of Hematology Calibrator (H Cal)
3. Printouts
4. Traceability Document

### L-J QC Table

Editor: admin      QC Mode: Whole Blood-CBC+DIFF      Lot No.: EH2205  
 Existing / Total: 15/500      Level: High      Date Range: 2022/05/11-2022/05/25  
 QC Sample ID:      Exp. Date: 2022/07/10      Print Time: 2022/05/28 17:06:35

Item	Ques	Time	WBC	Neut%	Lymph%	Mon%	Eosin%	Bas%	Neut	Lymph	Mon	RBC	HGB	HCT	MCV	MCH	MCHC	RDW-CV	RDW-SD	PLT	MPV	PDW-SD	PLV-CV	PCT	PLV-DC	PLV-CC		
Target	/	/	17.80	62.8	20.1	5.5	11.8	2.0	11.25	3.61	0.99	212	0.36	6.41	17.0	52.9	97.7	31.8	32.5	14.1	56.1	489	9.1	12.4	17.0	0.447	36.6	150
Unit (M)	/	/	2.50	7.0	6.0	5.5	7.0	2.0	1.40	1.10	0.99	1.30	0.30	0.8	4.0	6.0	2.8	3.0	3.0	12.0	60	3.0	3.0	0.200	8.0	35		
1	2022/05/25	15:42:21	18.11	62.8	19.8	5.3	12.5	0.6	11.34	3.55	0.96	2.26	0.10	5.22	16.5	50.4	96.5	31.5	32.7	14.3	55.8	508	11.0	13.2	17.6	0.456	32.7	155
14	2022/05/24	15:38:23	17.86	64.2	19.8	5.8	11.4	0.4	11.46	3.36	1.00	2.64	0.07	5.22	16.4	50.3	96.4	31.4	32.5	14.4	52.1	487	11.1	13.0	17.8	0.471	33.4	153
13	2022/05/23	15:31:48	18.19	63.4	19.6	5.3	11.7	0.5	11.53	3.57	0.99	2.13	0.08	5.22	16.4	50.2	96.3	31.5	32.7	14.3	55.9	490	11.0	12.7	17.3	0.485	33.3	159
12	2022/05/22	15:28:26	18.08	63.7	19.1	5.5	11.7	0.5	11.52	3.45	0.99	2.12	0.09	5.20	16.5	50.7	96.3	31.4	32.6	15.3	52.0	505	11.1	12.6	18.1	0.444	34.0	168
11	2022/05/21	15:25:36	18.17	63.8	20.0	4.9	11.5	0.6	11.56	3.80	0.89	2.09	0.11	5.20	16.6	50.2	96.5	31.8	33.0	15.1	55.1	499	11.1	15.0	18.8	0.455	33.2	160
9	2022/05/20	15:22:22	18.23	63.6	19.7	5.1	11.6	0.6	11.60	3.56	0.93	2.11	0.11	5.25	16.6	50.9	96.8	31.5	32.6	15.3	56.8	501	11.0	15.1	19.0	0.457	33.2	159
8	2022/05/18	15:19:09	18.18	63.0	20.0	5.1	11.9	0.4	11.45	3.64	0.93	2.16	0.07	5.28	16.8	51.0	97.1	31.5	32.5	14.1	55.2	497	11.0	14.7	18.9	0.384	33.0	167
7	2022/05/17	15:10:02	17.24	62.8	20.0	5.3	11.9	0.6	10.83	3.43	0.91	2.05	0.10	5.24	16.5	50.8	97.0	31.5	32.5	15.1	57.6	497	11.0	13.6	19.2	0.459	33.1	158
6	2022/05/16	15:07:18	16.82	63.1	20.0	4.9	12.0	0.4	11.38	3.60	0.88	2.16	0.08	5.25	16.6	50.9	97.1	31.7	32.6	15.1	61.3	508	10.3	13.9	18.1	0.508	30.5	148
5	2022/05/15	14:55:34	18.20	63.1	20.7	4.6	11.4	0.4	11.42	3.78	0.88	2.06	0.08	5.18	16.6	50.2	96.9	32.0	33.1	15.8	60.3	490	11.0	15.0	18.8	0.452	33.2	159
4	2022/05/14	14:34:20	18.92	62.6	20.6	4.7	12.1	0.5	11.84	3.90	0.89	2.29	0.09	5.43	17.4	53.0	97.6	32.0	32.8	15.0	61.8	495	11.1	14.8	19.0	0.442	33.7	160
3	2022/05/13	13:17:30	17.89	63.0	18.8	5.1	12.1	0.5	11.28	3.54	0.91	2.16	0.09	5.26	16.6	51.8	96.5	31.6	32.1	15.2	63.0	491	10.6	15.1	19.0	0.446	32.1	161
2	2022/05/12	13:15:14	18.03	63.0	19.9	4.7	12.4	0.5	11.35	3.90	0.85	2.24	0.09	5.22	16.6	51.3	96.4	31.6	32.3	15.0	58.0	489	11.0	14.0	18.4	0.482	32.9	160
1	2022/05/11	13:13:22	18.01	63.2	19.4	5.4	12.0	0.5	11.39	3.49	0.97	2.16	0.09	5.26	16.7	51.5	96.2	31.6	32.3	14.7	56.9	487	11.0	13.7	18.2	0.482	33.3	172

# L-J QC Graph

Lot No.: EH2205  
 Exp. Date: 2022/07/10  
 Current Page/All: 15/15  
 Print Time: 2022/05/26 17:

tor: admin Level: High  
 No.: 12 QC Mode: Whole Blood-CBC+DIFF  
 te Range: 2022/05/11 - 2022/05/25  
 Sample ID:

