



EURO CALIBRATION TECHNIQUE

No. 387, IInd Floor, TIC Road, Varadharajapuram,
Ambattur, Chennai-600 053. Tamil Nadu

Tel: 044-48521288 Mobile: 9486195184, 9176032184
E-mail: eurocalibration@gmail.com web: www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2022/0224- 01
ULR Number : CC2519 22 0 00001985F

Date of Issue : 7-Mar-2022

Pages : 1 of 1

Customer Name and Address

SRF Number : 2022/0224

Receipt Date : 4-Mar-2022

M/s. Government Hospital Krishnapuram,
ICTC Laboratory,

Date of Calibration : 4-Mar-2022

Attur Main Road, Krishnapuram,

Suggested calibration Due Date : 3-Mar-2023

Perambalur Dist - 621 116.

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Refrigerator

Make : Videocon

Range : 2 to 8 °C

Model : REFVZ293MESN-HFK

Resolution : NA

S.No : 210416860175000982

Accuracy : As Per Manual

ID No. : --

Location : Laboratory

Environmental Conditions

Temperature 26.4 °C

Relative Humidity 62 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor	1704	ECT 2021/0826-05	05.11.2022
2	Digital RTD Calibrator	12030078887	CT21-401-1	16.11.2022

Calibration Results

Parameter - Thermal / Temperature

S.No	Test Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Uncertainty ±°C
		1	2	3		
1	2 to 8	6.3	6.4	6.4	6.37	0.32

Checked By



Authorized Signatory
K.ABDUL VAJITH / R.SATHYAKALA

REMARKS: 1. Calibration certificate traceable to National / international standards which realize the units of measurement according to the international system of units (SI).
2. This certificate shall not be reproduced except in full without return approval of Euro Calibration Technique. 3. The Calibration results reported in the certificate as valid at the time of and under the stated condition of measurement for the particular equipment. 4. Next Calibration due date mentioned as per the customer requirement. 5. The measurement confidence level is 95% for a coverage factor k=2. 6. Calibration certificate issued for weights & measure parameter i.e mass, balance, volumetric equipments are for scientific or industrial purpose only.