

CALIBRATION CERTIFICATE

Page 1 of 1

EQN/FT/7.8/01			
CRF No.	EQN/CRF/2202124	Date of Calibration	24-Feb-22
Certificate No.	2021-22/EQN/2202124-28	Recom. Due Date	23-Feb-23
ULR No.	CC2760221000001471		

Customer Details:	Calibrated at	Onsite
M/s. BLOCK MEDICAL OFFICER	Date of Receipt	24-Feb-22
Government Primary Health Centre	Cond. On Receipt	Satisfactory
Nallampalayam,	Date of Issue	27-Feb-22
Coimbatore		

DETAILS OF UUC :		Identification No.	..
Description	Refrigerator	Serial No.	7187657204145490
Range	2-8°C	Make	Whirlpool
Least Count	---	Model / Type	DC 260G
Cal./Working Range	2-8°C	Accuracy	As Per Manual
Location	ICTC		

DETAILS OF STANDARD USED :			
Master Name	Lab ID No. / Sr No.	Valid upto	Certificate Number
Portable Calibrator	INS/ET/10	11-May-22	2020-21/EQN/MASTER-02
RTD	INS/TM/29	14-Aug-22	2021-22/EQN/210820-46

CP No.	EQN/CP/TL-03
Environmental Details	Temperature: 25±5°C Relative Humidity: 45-70 % RH

THERMAL CALIBRATION - (Temperature)

CALIBRATION RESULT			
Sl.No.	TEST READING	STD READING	UNCERTAINTY
	(°C)	(°C)	± (°C)
1	2.8	4.44	0.60

Remark :

1. Reading given was the average of 3 Measurements.
2. The reported Expanded Uncertainty is calculated at 95 % C.I. Approx. with coverage factor $k=2$
3. Temperature scale: ITS 90

Issued by :		Authorized by :	
Engineer - Calibration	Sangeetha K P		V. Rajkarthick



EQUATION CALIBRATION SERVICES PRIVATE LIMITED