



BIO-CHEMISTRY-01

ALL METHOD REPORT

Lab Code: 2246

Cycle - 11/2022
Round No - 1

Date: 21/02/2022

Parameters	Units	No.of. Parti- cipants	Group Mean	Standard deviation (SD)	Uncertai nty of Assign Values	Range (± 2 SD)	Your Value	Standard Deviation Index (SDI)
Albumin	g/dL	140	4.0	0.3	0.03	3.5-4.6	4.3	1.0
Alkaline Phosphatase	U/L	132	172.8	41.1	4.47	90.6-255.1	218	1.1
Bilirubin Total	mg/dL	141	1.5	0.2	0.02	1.1-1.9	1.6	0.5
Calcium(Total)	mg/dL	122	8.9	0.7	0.08	7.5-10.2	7.8	-1.6
Cholesterol (Total)	mg/dL	140	178.4	13.6	1.44	151.2-205.6	193	1.1
Creatinine	mg/dL	143	1.3	0.1	0.01	1.1-1.6	1.4	1.0
Glucose	mg/dL	142	111.1	6.9	0.72	97.3-124.9	109	-0.3
HDL	mg/dL	133	64.8	16.7	1.81	31.4-98.2	71.0	0.4
Potassium	mEq/L	125	3.8	0.2	0.02	3.4-4.2	4.0	1.0
Protein (Total)	g/dl	139	5.8	0.4	0.04	5.0-6.6	5.6	-0.5
Sodium	mEq/L	125	137.6	3.8	0.42	130.0-145.3	140	0.6
SGPT/ALT	U/L	141	31.7	4.8	0.51	22.1-41.2	41	1.9
SGOT/AST	U/L	141	33.0	5.1	0.54	22.8-43.1	39	1.2
Triglyceride	mg/dL	140	90.5	8.1	0.86	74.3-106.6	96	0.7
Urea	mg/dL	143	43.9	4.5	0.47	35.0-52.8	36	-1.8
Uric Acid	mg/dl	142	5.6	0.6	0.06	4.5-6.7	6.1	0.8

Interpretation of SDI:

SDI Value(+/-)	0 - 0.5	0.6 - 0.9	1.0 - 2.0	2.1 - 2.9	≥ 3
Interpretation	Excellent Performance	Good Performance	Acceptable Performance	Marginal Performance Need Improvement	Unacceptable Performance Needs Urgent action

Legends	(*) Excluded From Group Mean	(.) Not Reported	(#)Late Result Submission	(S)Reported in other Unit
---------	---------------------------------	------------------	------------------------------	------------------------------

Chief Coordinator


Dr.Sanjay Mehrotra

Checked By:

Programme Director

Dr.Bandana Mehrotra

Page 1 of 7



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence