

RAPID LABORATORY (607)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 02

Report Date : 23/05/2022

Instrument : Erba EM 200

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	5.80	5.51	1.10	0.79

Legend @ : z score \leq 2.0 - Acceptable	Total Parameters	1
○ : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score \geq 3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score HbA1c	100.00 %
* : Not considered for evaluation.		

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 IT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416
 Commercial Building - 1A,
 Kohnoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report

RAPID LABORATORY (607)


Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : Erba EM 200

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	6.60	6.36	0.63	0.79
Legend @ : z score \leq 2.0 - Acceptable					
! : 2.0 < z score < 3.0 - Warning					
X : z score \geq 3 - Unacceptable					
# : Not Evaluated					
⌚ : Delayed Result Entry					
* : Not considered for evaluation.					
				Total Parameters	1
				Not Evaluated Parameters	0
				Evaluated Parameters	1
				Outlier Parameters (X)	0
				EQAS Score HbA1c	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohnoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report

RAPID LABORATORY (607)
Outlier And Analyte Summary Report
Result Details For Cycle No 220104 and Sample No 02

Report Date : 23/05/20

Instrument : Erba EM 200

Analyte

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose					
@ Urea	mg/dl	303.00	288.48	0.75	0.99
@ Creatinine	mg/dl	21.80	20.79	0.83	0.73
@ Bilirubin Total	mg/dl	2.60	2.82	-1.01	-0.75
@ Total Protein	mg/dl	0.42	0.44	-0.40	0.26
@ Albumin	g/dL	9.00	8.41	0.79	1.05
@ AST	g/dL	5.09	4.96	0.46	-0.17
@ ALP	U/L	16.70	17.95	0.17	0.46
@ Uric Acid	U/L	34.00	43.18	-1.05	-0.48
@ Calcium	mg/dl	4.30	4.01	0.74	0.38
@ Cholesterol	mg/dl	7.50	7.55	-0.18	-0.69
@ Triglycerides	mg/dl	138.00	132.33	0.54	0.45
@ HDL	mg/dl	97.40	99.73	-0.25	-0.02
@ ALT	mg/dl	36.50	34.06	0.53	0.54
	U/L	5.80	8.96	-1.49	-0.32

Instrument : HDC- LYTE

Analyte

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Potassium	mmol/L	3.60	-	-0.33	-0.43

Legend @ : z score \leq 2.0 - Acceptable
 : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	15
Not Evaluated Parameters	0
Evaluated Parameters	15
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

DK Nigam

Puneet Kumar Nigam
 coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416
 Commercial Building - 1A,
 Kothinoor Mall, Kiroli Road, Kurta (W),
 Mumbai - 400070

RAPID LABORATORY (607)

Outlier And Analyte Summary Report

Result Details For Cycle No 220104 and Sample No 01

Report Date : 23/05/2022

Instrument : Erba EM 200


Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	76.40	70.95	0.92	0.99
@ Urea	mg/dl	23.00	22.52	0.68	0.73
@ Creatinine	mg/dl	0.81	0.80	-0.24	-0.75
@ Total Protein	g/dL	7.50	6.70	1.57	1.05
@ Albumin	g/dL	3.70	3.61	0.45	-0.17
@ Uric Acid	mg/dl	4.50	4.68	-0.14	0.38
@ ALP	U/L	45.00	54.09	-1.06	-0.48
@ Calcium	mg/dl	7.80	8.05	-0.47	-0.69
@ Cholesterol	mg/dl	145.00	142.68	0.37	0.45
@ Triglycerides	mg/dl	103.00	106.75	-0.20	-0.02
@ HDL	mg/dl	37.10	35.50	0.45	0.54
@ ALT	U/L	6.10	8.61	-0.81	-0.32
@ AST	U/L	17.10	16.87	0.20	0.46

Instrument : HDC- LYTE

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Sodium	mmol/L	133.00	-	0.76	1.17
@ Potassium	mmol/L	3.67	-	-0.63	-0.43

Legend @ : z score \leq 2.0 - Acceptable
 † : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	15
Not Evaluated Parameters	0
Evaluated Parameters	15
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr. Puneet Kumar Nigam
 coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416
 Commercial Building - 1A,
 Kashiroad Mall, Kiroli Road, Kurli (W),
 Mumbai - 400070

End of Report